

Benchmark Midplane Field for Medical FFA

S. Brooks

April 2024

Collider Accelerator Department
Brookhaven National Laboratory

U.S. Department of Energy
USDOE Office of Science (SC), Nuclear Physics (NP)

Notice: This technical note has been authored by employees of Brookhaven Science Associates, LLC under Contract No. DE-SC0012704 with the U.S. Department of Energy. The publisher by accepting the technical note for publication acknowledges that the United States Government retains a non-exclusive, paid-up, irrevocable, world-wide license to publish or reproduce the published form of this technical note, or allow others to do so, for United States Government purposes.

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or any third party's use or the results of such use of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof or its contractors or subcontractors. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

Benchmark Midplane Field for Medical FFA

Stephen Brooks, Dejan Trbojevic

April 2, 2024

1 Introduction

This is a lattice cell for a fixed-field medical accelerator, transmitting protons from 10–250 MeV. The goal is to have fixed tunes throughout this energy range. The cell is based on the lattice `Feb7_2024.ProtonAccelerator.f90` from Dejan Trbojevic and had the magnet field profiles re-optimised to attain fixed tunes with realistic magnet fringe fields.

1.1 Coordinate Convention

z is the forward longitudinal coordinate, x is transverse (orbit excursion) and y is vertical. The curvature of the machine is anticlockwise as seen from above (positive y side). x points towards the machine centre, so the high energy orbits are at more negative x .

2 Cell Geometry

The cell as drawn in Figure 1 has element start and end coordinates given in Table 1.

The cell starts at the beginning of the first half drift $H0$ at $(0, 0, 0)$. It ends at the end of the second half drift $H0$, with total cell rotation angle $0.09817477 \text{ rad} = 5.625^\circ = 360^\circ/64$.

The “Length” and “Angle” values define a sequence of circular arcs and straight lines that form the reference axis for the cell. However, the rectangular magnets are only aligned ‘on average’ on the curved parts, so the start and end coordinates for the magnets do not lie on the reference axis.




Figure 1: Closed orbits in Muon1 on top of the cell geometry and 1 cm grid.

Table 1: Element start and end positions in absolute coordinates.

Element name	Length (m)	Angle (rad)		
HO	0.0045	0		
F	0.212	0.0639198516350064		
O	0.009	0		
D	0.113611696486384	0.0342549187896746		
HO	0.0045	0		
Element name	Start x (m)	y	z	
HO	0	0	0	
F	-0.00112915456451474	0	0.00451804630577078	
O	0.00677319767087102	0	0.216355666412717	
D	0.00702461901507897	0	0.22536077394586	
HO	0.0165454801670591	0	0.33857051247452	
Element name	End x (m)	y	z	
HO	0	0	0.0045	
F	0.00564519630940407	0	0.216409783214592	
O	0.00734808467505134	0	0.225337286808416	
D	0.0162224641982652	0	0.338599535960023	
HO	0.0169865572985422	0	0.343048843744545	

3 Midplane ($y = 0$) Magnetic Field Model

The field of each magnet is defined in terms of local Cartesian coordinates z_{loc} and x_{loc} , where $(z_{loc}, x_{loc}) = (0, 0)$ is the start coordinate of the magnet and $(z_{loc}, x_{loc}) = (L, 0)$ is the end.

The midplane field is vertical and the product of three functions:

$$B_y(x, 0, z) = g(x_{loc})f(z_{loc})f(L - z_{loc}), \quad B_x(x, 0, z) = B_z(x, 0, z) = 0,$$

where g defines the body field and f are fringe field sigmoid-type functions.




Figure 2: Colour plot of the vertical field component in the cell, colour contours are 0.1 T.

Figure 2 shows the field in the cell.

3.1 Body Field Function

To avoid ill-conditioned optimisation with high-order polynomials, the body field function is defined using a Fourier series:

$$g(x) = c_0 + \sum_{n=1}^8 s_n \sin(nkx) + c_n \cos(nkx),$$

where k is a transverse scale factor and $c_0...c_8$ and $s_1...s_8$ are the coefficients defining the field (units Tesla).

The values for these coefficients of the two magnets are given in Table 2.

Table 2: Fourier coefficients for the magnet body fields.

Magnet F		
	$k = 60 \text{ (m}^{-1}\text{)}$	
n	$c_n \text{ (T)}$	$s_n \text{ (T)}$
0	-0.69802703489829	—
1	0.0756049884526512	1.73518865753323
2	0.305088336939702	-0.442262271952788
3	-0.139524713341402	0.0826714806226137
4	0.0276793694098833	-0.0208114430730033
5	-0.00196329427184875	0.0191643586396017
6	0.00341848475778073	-0.0112392171957729
7	-0.00264986640664568	0.00271130619879153
8	0.000542396377167934	-6.52682264979307E-05
Magnet D		
	$k = 60 \text{ (m}^{-1}\text{)}$	
n	$c_n \text{ (T)}$	$s_n \text{ (T)}$
0	0.989342248880683	—
1	-1.91337116068355	-2.46074568747532
2	0.397559249919083	0.493240962238742
3	-0.0885340223614818	-0.134280872283561
4	0.0116131686129637	0.0706611962587838
5	-0.010779760306873	0.00724018936522096
6	0.0207629971432377	-0.0480848747522068
7	-0.014783458990518	0.0306916335109993
8	0.00390264678161641	-0.00685949055711114

3.2 Fringe Field Function

It was found that functions of the form

$$f(z) = \frac{\frac{1}{a} \arctan(z/a) - \frac{1}{b} \arctan(z/b)}{\left(\frac{1}{a} - \frac{1}{b}\right) \pi} + \frac{1}{2}$$

are a good fit to real permanent magnet fringe fields.

Both magnets in this cell use the values $a = 0.008 \text{ m}$ and $b = 0.04 \text{ m}$.

The use of arc tangent is physically motivated from the field of two half-infinite current sheets with the beam emerging between them. This would have a long-range decay as $1/z$, which is too slow, so the difference of two arc tangents is used, cancelling the $1/z$ term and leaving the remainder decaying as $1/z^3$. This is the field of two current sheets that are infinite transversely but not vertically.

This fringe field is symmetrical so magnets can be split into longitudinal segments while keeping the total field the same.

Appendix A Muon1 Lattice Input File

```
{Dejan's medical fixed-tune NSFFA for LDRD, design 2024-Feb-07
Proper Fourier approximation (order 5), by least squares fit, using  $k=60m^{-1}$ , zero
  extended to order 8
But with TwoAtan fringe fields fitted to the 3D magnet model, increase FringeLimit to 50
  cm
Optimised values including max |Bmidl|}

#// Dejan/medicalfixedtune_LDRD_2024-Feb-07 - Composite cell of 17 elements:

{MatchScan AllowUnstable Egoal Estart Species}
Match 1 1.602176634e-012 4.005441585e-011 'Proton'
{Drift Length}
HO 0.0045
{Magnet3Rect AlignMode Angle FringeFunc FringeFunca FringeFuncb FringeLimit func3 func3c0
  func3s1 func3c1 func3s2 func3c2 func3s3 func3c3 func3s4 func3c4 func3s5 func3c5
  func3s6 func3c6 func3s7 func3c7 func3s8 func3c8 func3order func3k HalfHeight
  HalfWidth Length}
F 'Integral' 0.0639198516350064 'TwoAtan' 0.008 0.04 0.5 'Fourier' -0.69802703489829
  1.73518865753323 0.0756049884526512 -0.442262271952788 0.305088336939702
  0.0826714806226137 -0.139524713341402 -0.0208114430730033 0.0276793694098833
  0.0191643586396017 -0.00196329427184875 -0.0112392171957729 0.00341848475778073
  0.00271130619879153 -0.00264986640664568 -6.52682264979307E-05 0.000542396377167934 8
  60 0.01 0.07 0.212
{Drift Length}
O 0.009
{Magnet3Rect AlignMode Angle FringeFunc FringeFunca FringeFuncb FringeLimit func3 func3c0
  func3s1 func3c1 func3s2 func3c2 func3s3 func3c3 func3s4 func3c4 func3s5 func3c5
  func3s6 func3c6 func3s7 func3c7 func3s8 func3c8 func3order func3k HalfHeight
  HalfWidth Length}
D 'Integral' 0.0342549187896746 'TwoAtan' 0.008 0.04 0.5 'Fourier' 0.989342248880683
  -2.46074568747532 -1.91337116068355 0.493240962238742 0.397559249919083
  -0.134280872283561 -0.0885340223614818 0.0706611962587838 0.0116131686129637
  0.00724018936522096 -0.010779760306873 -0.0480848747522068 0.0207629971432377
  0.0306916335109993 -0.014783458990518 -0.00685949055711114 0.00390264678161641 8 60
  0.01 0.07 0.113611696486384
{Drift Length}
HO 0.0045
{Match-Aperture}
MatchEnd

Dejan_medicalfixedtune_LDRD_2024_Feb_07: HO,F,O,D,HO,Match,HO,F,O,D,HO,MatchEnd,HO,F,O,D,
HO;
```