

### BNL-90927-2010-TECH EIC/19;BNL-90927-2010-IR

## **R&D ERL:** Power Supplies

R. Lambiase,

January 2010

Collider Accelerator Department Brookhaven National Laboratory

### **U.S. Department of Energy**

USDOE Office of Science (SC)

Notice: This technical note has been authored by employees of Brookhaven Science Associates, LLC under Contract No.DE-AC02-98CH10886 with the U.S. Department of Energy. The publisher by accepting the technical note for publication acknowledges that the United States Government retains a non-exclusive, paid-up, irrevocable, world-wide license to publish or reproduce the published form of this technical note, or allow others to do so, for United States Government purposes.

### DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or any third party's use or the results of such use of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof or its contractors or subcontractors. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

C-A/AP/#373 January 2010

# **R&D ERL: Power Supplies**

R. Lambiase



### Collider-Accelerator Department Brookhaven National Laboratory Upton, NY 11973

Notice: This document has been authorized by employees of Brookhaven Science Associates, LLC under Contract No. DE-AC02-98CH10886 with the U.S. Department of Energy. The United States Government retains a non-exclusive, paid-up, irrevocable, world-wide license to publish or reproduce the published form of this document, or allow others to do so, for United States Government purposes.

### R&D ERL – Power Supplies R. Lambiase December 18, 2009

A magnet power supply system has been developed to meet the field requirements of the ERL in a compact and cost effective fashion.

### **Design Considerations**

The magnet assemblies used in the ERL consist of one or more windings on a common core. Each of the windings represents a separate magnet load for the power supply. As the ERL is operated in a DC fashion, interaction between the windings is not a concern.

Appendix A lists all the magnet assemblies by sector. Each coil is listed by name and model. The model corresponds to a set of electrical and magnet parameters as established by the magnet subsystem.

Some of the coils are connected in series. The connection scheme, plus cabling provides the electrical load characteristics. The load information, plus the operating current, and the stability define the power supply requirements.

#### Power Supply Listing

All of the magnet power supply requirements for ERL can be satisfied by five different models. The capsule specifications and quantities are shown in Table 1 below.

Model	Voltage, Volts	Current, Amps	Precision,	Quantity
UD320A35V, IE Power	35	320	100	1
Shim Amplifier 892, Danfysik	15	10	100	34
BOP 50-20GL, Kepco	50	20	100	5
MCOR12 / 2A, BiRa	25	2	1000	32
MCOR12 / 6A, BiRa	25	6	1000	6

With the exception of the UD320A35V unit, all models are bipolar, even though not all loads require bipolar operation. But, by using standard off-the-shelf units, development costs were minimized.

Appendix B lists each power supply, by model, with its magnet load.

### UD320A35V, IE Power

Ratings: 320A, 35V, 100ppm Qty: 1

This supply is used to power the six main dipoles in series. It can be seen as the cabinet on the left in Figure 1.

This supply is a thyristor controlled supply using the same design, but slightly different rating, as the 77 medium range power supplies built for the SNS at ORNL.

This supply is controlled by a standard BNL Power Supply Interface (PSI).

BOP 50-20GL, Kepco

Ratings: 50V, 20A, 100ppm Qty: 5

This supply is used where higher voltage is needed, typically where there are more than one coil in series as a load. All five units can be seen mounted in the cabinet on the right in Figure 1.

These switch mode supplies are a high precision version of the standard Kepco



Figure 1. IE Power & Kepco Supplies

high power BOP. The extra precision is obtained by closing the current loop around a zero flux current sensor (ZFCT) and adding additional output filtering to reduce the output ripple to a level consistent with the higher precision.

Control of these supplies will be by RS-232.

### Shim Amplifier 892, Danfysik

Ratings: 10A, 15V, 100ppm Qty: 34

All 34 supplies, plus two spares, are shown in Figure 2. Each crate consists of a bulk power supply and six regulators. Each regulator has a front end switching preregulator followed by a linear H bridge. The high stability is maintained by an on-board ZFCT.

These power supplies are scaled down versions of the 20A, 70V low field correctors designed for the ORNL SNS. By reducing the power, these were able to be packaged in a compact configuration, and used to power shim windings on MRI machines. It's also very useful for us to have this many high precision power supplies in a small volume.

Each channel can be individually manually controlled by the controller seen at the top of the left cabinet. It has the capability to address and control up to 256 regulators.

In operation, these supplies will be controlled by a RS-485 line.



Figure 2. Danfysik Power Supplies

#### MCOR12 / 2A, BiRa

Rating: 2A, 25V, 1000ppm Qty: 36

### MCOR12 / 6A, BiRa

Rating: 2A, 25V, 1000ppm Qty: 36

These magnet power supplies are shown in Figure 3. The 32 regulators rated at 2A each in two crates in the left cabinet, and the six regulators rated at 6A each are in the crate in the right cabinet.

This magnet corrector power supply system was designed at SLAC, and built to their specification by BiRa.

Each crate contains one analog interface board and up to 16 MCOR12 regulator cards that can be configured as 2A, 6A, or 12A by means of a programming daughter board.

These regulators are powered by a commercial unipolar power supply. In our application, these are Genesys 30V, 50A units.

Completing the system is a 2U blower assembly which goes between the



Figure 3. BiRa Power Supplies

bulk power supply and the crate. It draws air from the front panel through a filter and directs it up between the cards.

Each regulator accepts (via the interface card) one analog current set point, and returns one analog current read back. These signals will come from VME DACs and ADCs in the control crate. In addition, each crate has a group interlock, and group status.

An RS-232 digital interface, which includes the ADC and DAC functions is under development at SLAC, but is not completed at this time.

### Rack Layout

The magnet power supplies are all located in a room above the transmitter water room. The arrangement of racks is shown in Figure 4.




Figure 4. Power Supply Rack Layout

### **Construction Status**

All of the magnet power supplies have been delivered, and are nearly fully installed mechanically. Work will continue with installing cables for AC power to the equipment, DC cables from the equipment to the magnets, and control cables from the control system to the power supplies.

This work should be completed by February 2010.

### **Commissioning Plans**

All of the power supplies have been tested at the factory, and there are no technical complexities to commissioning these supplies. The control interface design will be tested prior to the actual magnet loads being connected and the final system testing will proceed as the magnet loads are connected.

### Appendix A

### Magnet Configuration by Sector

Sector: 1			
	h Temp SC solenoid	Madal 10	
Assembly Name: El		Model: 10	
Coil Name: El	LS1.1 COI	Name: 10	0510H15
Sector: 2			
0	coils around 6-way cross		
Assembly Name: El		Model: 5D	D10
Coil Name: El	LC2.1CH Coi	Name: 5D	DH10
Coil Name: El	LC2.1CV Coi	Name: 5D	DV10
Coil Name: El	LS2.1 Coi	Name: 11	S10
Sector: 3			
-	degree combined function		
Assembly Name: El		Model: 70	
Coil Name: E		Name: 70	
Coil Name: El		Name: 70	
Coil Name: El		Name: 70	CQ15
Coil Name: El	LD3.1X Coi	Name: 70	CX15
0	degree combined function		
Assembly Name: E		Model: 70	
Coil Name: El		Name: 70	
Coil Name: El		Name: 70	
Coil Name: El	LD3.2Q Coi	Name: 70	CQ30
Coil Name: El	LD3.2X Coi	Name: 70	CX30
	degree combined function		
Assembly Name: E		Model: 70	
Coil Name: El		Name: 70	
Coil Name: El	LD3.3CH Coi	Name: 70	CDHT15
Coil Name: El	LD3.3Q Coi	Name: 70	CQ15
Coil Name: El	LD3.3X Coi	Name: 70	CX15
-	degree combined function		_
Assembly Name: El		Model: 70	
Coil Name: El		Name: 70	
Coil Name: El	LD3.4CH Coi	Name: 70	CDHT30
Coil Name: El	LD3.4Q Coi	Name: 70	CQ30
Coil Name: El	LD3.4X Coi	Name: 70	CX30

### Sector:4

Sector:4			
Magnet: 4 - 1 : Injection line sole			
Assembly Name: ELS4.1	Assembly Model: 11S10		
Coil Name: ELS4.1	Coil Name: 11S10		
Magnet: 4 - 2 : Injection line sole	noid		
Assembly Name: ELS4.2	Assembly Model: 11S10		
Coil Name: ELS4.2	Coil Name: 11S10		
Magnet: 4 - 3 : Extraction line sol	enoid		
Assembly Name: ELS4.3	Assembly Model: 11S10		
Coil Name: ELS4.3	Coil Name: 11S10		
Magnet: 4 - 4 : Extraction line sol	enoid		
Assembly Name: ELS4.4	Assembly Model: 11S10		
Coil Name: ELS4.4	Coil Name: 11S10		
Sector: 5			
Magnet: 5 - 1 : 1st Extraction dipe	ole		
Assembly Name: ELD5.1	Assembly Model: 6D20		
Coil Name: ELD5.1	Coil Name: 6D20		
Coil Name: ELD5.1CH	Coil Name: 6DT20		
Coil Name: ELS5.1	Coil Name: 11S10		
Magnet: 5 - 2 : Compensating chicane			
Assembly Name: ELD5.2	Assembly Model: 6D40		
Coil Name: ELD5.2	Coil Name: 6D40		
Coil Name: ELD5.2CH	Coil Name: 6DT40		
Coil Name: ELS5.2	Coil Name: 11S10		
Magnet: 5 - 3 : Compensating ch	icane		
Assembly Name: ELD5.3	Assembly Model: 6D20		
Coil Name: ELD5.3	Coil Name: 6D20		
Coil Name: ELD5.3CH	Coil Name: 6DT20		
Sector: 6			
Magnet: 6 - 1 : 60 degree dipole,	20 cm radius		
Assembly Name: ELD6.1	Assembly Model: 3D60		
Coil Name: ELD6.1	Coil Name: 3D60		
Coil Name: ELD6.1CH	Coil Name: 3DT60		
Coil Name: ELQ6.1	Coil Name: 6Q12		
Magnet: 6 - 2 : Regular quadrupo	ble		
Assembly Name: ELQ6.2	Assembly Model: 6Q12		

	Coil Name: ELQ6.2	Coil Name:	6Q12
	Coil Name: ELQ6.2CV	Coil Name:	6QTV12
Magnet:	6 - 3 : Regular quadrupol	е	
Asse	mbly Name: ELQ6.3	Assembly Model:	6Q12
	Coil Name: ELQ6.3	Coil Name:	6Q12

Sector: 7		
Magnet:	7 - 1 : 60 degree dipo	le, 20 cm radius
Asse	mbly Name: ELD7.1	Assembly Model: 3D60
	Coil Name: ELD7.1	Coil Name: 3D60
	Coil Name: ELD7.1CH	Coil Name: 3DT60
	Coil Name: ELQ7.1	Coil Name: 6Q12
	Coil Name: ELQ7.1CH	Coil Name: 6QTH12
Magnet:	7 - 2 : Regular quadru	ipole
Asse	mbly Name: ELQ7.2	Assembly Model: 6Q12
	Coil Name: ELQ7.2	Coil Name: 6Q12
	Coil Name: ELQ7.2CV	Coil Name: 6QTV12
Magnet:	7 - 3 : Regular quadru	ipole
Asse	mbly Name: ELQ7.3	Assembly Model: 6Q12
	Coil Name: ELQ7.3	Coil Name: 6Q12

#### Sector: 8

Magnet: 8 - 1 : 60 degree dipole, 20 cm radius			
Assembly Name:	ELD8.1	Assembly Model:	3D60
Coil Name:	ELD8.1	Coil Name:	3D60
Coil Name:	ELD8.1CH	Coil Name:	3DT60
Coil Name:	ELQ8.1	Coil Name:	6Q12
Magnet: 8 - 2 : R	egular quadrupole	Ð	
Assembly Name:	ELQ8.2	Assembly Model:	6Q12
Coil Name:	ELQ8.2	Coil Name:	6Q12
Coil Name:	ELQ8.2CV	Coil Name:	6QTV12
Magnet: 8 - 3 : R	egular quadrupole	Э	
Assembly Name:	ELQ8.3	Assembly Model:	6Q12
Coil Name:	ELQ8.3	Coil Name:	6Q12
Coil Name:	ELQ8.3CH	Coil Name:	6QTH12

### Sector: 9

Magnet: 9 - 1 : Regular quadru	apole
Assembly Name: ELQ9.1	Assembly Model: 6Q12
Coil Name: ELQ9.1	Coil Name: 6Q12
Magnet: 9 - 2 : Regular quadru	lode
Assembly Name: ELQ9.2	Assembly Model: 6Q12
Coil Name: ELQ9.2	Coil Name: 6Q12
Coil Name: ELQ9.2CV	Coil Name: 6QTV12

### Sector: 10

Magnet: 10 - 1 : Regular quadru	pole
Assembly Name: ELQ10.1	Assembly Model: 6Q12
Coil Name: ELQ10.1	Coil Name: 6Q12
Magnet: 10 - 2 : Regular quadru	pole
Assembly Name: ELQ10.2	Assembly Model: 6Q12
Coil Name: ELQ10.2	Coil Name: 6Q12

#### Sector: 11 11 - 1 : Regular quadrupole Magnet: Assembly Name: ELQ11.1 Assembly Model: 6Q12 Coil Name: ELQ11.1 Coil Name: 6Q12 11 - 2 : Regular quadrupole Magnet: Assembly Name: ELQ11.2 Assembly Model: 6Q12 Coil Name: ELQ11.2 Coil Name: 6Q12 Coil Name: ELQ11.2CH Coil Name: 6QTH12 Sector: 12 12 - 1 : 60 degree dipole, 20 cm radius Magnet: Assembly Name: ELD12.1 Assembly Model: 3D60 Coil Name: ELD12.1 Coil Name: 3D60 Coil Name: ELD12.1CH Coil Name: 3DT60 Coil Name: ELQ12.1 Coil Name: 6Q12 Magnet: 12 - 2 : Regular quadrupole Assembly Name: ELQ12.2 Assembly Model: 6Q12 Coil Name: ELQ12.2 Coil Name: 6Q12 Coil Name: ELQ12.2CV Coil Name: 6QTV12 Sector: 13 13 - 1 : 60 degree dipole, 20 cm radius Magnet: Assembly Name: ELD13.1 Assembly Model: 3D60 Coil Name: ELD13.1 Coil Name: 3D60 Coil Name: ELD13.1CH Coil Name: 3DT60 Coil Name: ELQ13.1 Coil Name: 6Q12 Coil Name: 6QTH12 Coil Name: ELQ13.1CH Magnet: 13 - 2 : Regular quadrupole Assembly Name: ELQ13.2 Assembly Model: 6Q12 Coil Name: ELQ13.2 Coil Name: 6Q12 Coil Name: ELQ13.2CV Coil Name: 6QTV12 Magnet: 13 - 3 : Regular quadrupole Assembly Name: ELQ13.3 Assembly Model: 6Q12 Coil Name: ELQ13.3 Coil Name: 6Q12

#### Sector: 14

Sector:	14			
Magne	et: 14 - 1 : 60	) degree dipole, 2	0 cm radius	
A	ssembly Name:	ELD14.1	Assembly Model:	3D60
	Coil Name:	ELD14.1	Coil Name:	3D60
	Coil Name:	ELD14.1CH	Coil Name:	3DT60
	Coil Name:	ELQ14.1	Coil Name:	6Q12
Magne	et: 14 - 2 : R	egular quadrupole	Э	
A	ssembly Name:	ELQ14.2	Assembly Model:	6Q12
	Coil Name:	ELQ14.2	Coil Name:	6Q12
	Coil Name:	ELQ14.2CV	Coil Name:	6QTV12
Magne	et: 14 - 3 : R	egular quadrupole	Э	
A	ssembly Name:	ELQ14.3	Assembly Model:	6Q12
	Coil Name:	ELQ14.3	Coil Name:	6Q12
	Coil Name:	ELQ14.3CH	Coil Name:	6QTH12

#### Sector:15

Magnet: 15 - 1 : Small compensating dipole			
mbly Name: ELD15.1	Assembly Model: 3D2		
Coil Name: ELD15.1	Coil Name: 3D2		
Coil Name: ELQ15.1	Coil Name: 3Q12		
15 - 2 : Small compen	sating dipole		
mbly Name: ELD15.2	Assembly Model: 3D2		
Coil Name: ELD15.2	Coil Name: 3D2		
Coil Name: ELQ15.2	Coil Name: 3Q12		
	mbly Name: ELD15.1 Coil Name: ELD15.1 Coil Name: ELQ15.1 15 - 2 : Small compen mbly Name: ELD15.2 Coil Name: ELD15.2		

#### Sector: 16

Magnet:	agnet: 16 - 1 : Corrector in extraction line		
Asse	mbly Name: ELC16.1	Assembly Model: 5D10	
	Coil Name: ELC16.1C	H Coil Name: 5DH10	
	Coil Name: ELC16.1C	V Coil Name: 5DV10	
	Coil Name: ELD16.1	Coil Name: 6D30	
	Coil Name: ELS16.1	Coil Name: 10S10	
Magnet:	16 - 2 : Driving to Be	am Dump Solenoid	
Asse	mbly Name: ELS16.2	Assembly Model: 10S10	
	Coil Name: ELS16.2	Coil Name: 10S10	
Magnet:	16 - 3 : Driving to Be	am Dump Solenoid	
Asse	mbly Name: ELS16.3	Assembly Model: 10S10	
	Coil Name: ELS16.3	Coil Name: 10S10	
Magnet:	16 - 4 : Driving to Be	am Dump Solenoid	
Asse	mbly Name: ELS16.4	Assembly Model: 10S10	
	Coil Name: ELS16.4	Coil Name: 10S10	

### Appendix B Magnet Circuits by Power Supply Model

UD320A35V, IE Power		20	. PS_ELQ15.1
1. PS ELDMain		21	<i>ELQ15.1</i> . PS_ELQ15.2
<i>ELD8.1</i>	3D60		ELQ15.2
ELD12.1	3D60	22	. PS_ELQ6.1
ELD12.1 ELD14.1	3D60	22	ELQ6.1
ELD6.1	3D60	22	. PS_ELQ6.2
ELD0.1 ELD13.1	3D60	23	
		0.4	ELQ6.2
ELD7.1	3D60	24	. PS_ELQ6.3 <i>ELQ6.3</i>
Shim Amplifier 892, Dan	fysik	25	. PS_ELQ7.1 <i>ELQ7.1</i>
		26	
1 . PS_ELD15.1	202	20	. PS_ELQ7.2
ELD15.1	3D2	07	ELQ7.2
2 . PS_ELD15.2	000	27	. PS_ELQ7.3
ELD15.2	3D2	~~	ELQ7.3
3 . PS_ELD16.1		28	. PS_ELQ8.1
ELD16.1	6D30		ELQ8.1
4 . PS_ELD3.1		29	. PS_ELQ8.2
ELD3.1	7CD15		ELQ8.2
5 . PS_ELD3.2		30	. PS_ELQ8.3
ELD3.2	7CD30		ELQ8.3
6 . PS_ELD3.3		31	. PS_ELQ9.1
ELD3.3	7CD15		ELQ9.1
7 . PS_ELD3.4		32	. PS_ELQ9.2
ELD3.4	7CD30		ELQ9.2
8 . PS_ELQ10.1		33	. PS_ELS1.1
ELQ10.1	6Q12		ELS1.1
9. PS_ELQ10.2		34	. PS_ELS2.1
ELQ10.2	6Q12		ELS2.1
10 . PS_ELQ11.1			
ELQ11.1	6Q12		
11 . PS_ELQ11.2			
ELQ11.2	6Q12		
12 . PS_ELQ12.1			
ELQ12.1	6Q12		
13 . PS_ELQ12.2			
<i>ELQ12.2</i>	6Q12		
14 . PS_ELQ13.1	0412		
ELQ13.1	6Q12		
15 . PS_ELQ13.2	0012		
ELQ13.2	6Q12		
16 . PS_ELQ13.3	0012		
	6012		
ELQ13.3	6Q12		
17 . PS_ELQ14.1	6012		
ELQ14.1	6Q12		
18 . PS_ELQ14.2	0040		
ELQ14.2	6Q12		
19 . PS_ELQ14.3			
ELQ14.3	6Q12		

20	. PS_ELQ15.1	
	ELQ15.1	3Q12
21	. PS_ELQ15.2	_
	ELQ15.2	3Q12
	. PS_ELQ6.1	
	ELQ6.1	6Q12
23	. PS_ELQ6.2	0010
74	<i>ELQ6.2</i> . PS ELQ6.3	6Q12
	<i>ELQ6.3</i>	6Q12
	. PS_ELQ7.1	UQ12
_0	ELQ7.1	6Q12
26	. PS_ELQ7.2	0G12
	ELQ7.2	6Q12
27	PS_ELQ7.3	
	ELQ7.3	6Q12
28	. PS_ELQ8.1	
	ELQ8.1	6Q12
29	. PS_ELQ8.2	
	ELQ8.2	6Q12
30	. PS_ELQ8.3	0040
24	<i>ELQ8.3</i> . PS ELQ9.1	6Q12
51	<i>ELQ9.1</i>	6Q12
22	. PS_ELQ9.2	UQ12
2	ELQ9.2	6Q12
33	. PS ELS1.1	0Q12
	ELS1.1	10S10HTS
34	. PS_ELS2.1	
	ELS2.1	11S10

### MCOR12/6A, BiRa

1 . PS_EL	D12.1CH	
ELD12.	1CH	3DT60
2 . PS_EL	D13.1CH	
ELD13.	1CH	3DT60
3 . PS_EL	D14.1CH	
ELD14.	1CH	3DT60
4.PS_EL	D6.1CH	
ELD6.1	CH	3DT60
5 . PS_EL	D7.1CH	
ELD7.1	CH	3DT60
6 . PS_EL	D8.1CH	
ELD8.1	CH	3DT60

### MCOR12/2A, BiRa

1	PS_ELC16.1CH		
_	ELC16.1CH	5DH10	
2	PS_ELC16.1CV	<b>FD</b> ) (4.0	
~	ELC16.1CV	5DV10	
3	. PS_ELC2.1CH	<b>FDU40</b>	
4	ELC2.1CH	5DH10	
4	PS_ELC2.1CV ELC2.1CV	5DV10	
5	. PS ELD3.1CH	50110	
5	ELD3.1CH	7CDHT15	В
6	. PS_ELD3.1Q	7001115	D
0	ELD3.1Q	7CQ15	
7	. PS_ELD3.1X	70010	
,	ELD3.1X	7CX15	
8	PS_ELD3.2CH	10/10	
0	ELD3.2CH	7CDHT30	
9	. PS_ELD3.2Q	10211100	
•	ELD3.2Q	7CQ30	
10	PS_ELD3.2X		
	ELD3.2X	7CX30	
11	. PS_ELD3.3CH		
	ELD3.3CH	7CDHT15	
12	. PS_ELD3.3Q		
	ELD3.3Q	7CQ15	
13	. PS_ELD3.3X		
	ELD3.3X	7CX15	
14	. PS_ELD3.4CH		
	ELD3.4CH	7CDHT30	
15	. PS_ELD3.4Q		
	ELD3.4Q	7CQ30	
16	. PS_ELD3.4X		
	ELD3.4X	7CX30	
17	. PS_ELD5.1CH		
	ELD5.1CH	6DT20	
18	PS_ELD5.2CH	0 <b>DT</b> (0	
	ELD5.2CH	6DT40	

	. PS_ELD5.3CH ELD5.3CH	6DT20
20	. PS_ELQ10.2CH ELQ10.2CH	6QTH12
21	. PS_ELQ11.2CH <i>ELQ11.2CH</i>	6QTH12
22	. PS_ELQ12.2CV ELQ12.2CV	6QTV12
23	. PS_ELQ13.1CH ELQ13.1CH	6QTH12
24	. PS_ELQ13.2CV	
25	<i>ELQ13.2CV</i> . PS_ELQ14.2CV	6QTV12
26	<i>ELQ14.2CV</i> . PS_ELQ14.3CH	6QTV12
27	<i>ELQ14.3CH</i> . PS_ELQ6.2CV	6QTH12
	<i>ELQ6.2CV</i> . PS_ELQ7.1CH	6QTV12
	<i>ELQ7.1CH</i> . PS_ELQ7.2CV	6QTH12
	ELQ7.2CV	6QTV12
	. PS_ELQ8.2CV <i>ELQ8.2CV</i>	6QTV12
31	. PS_ELQ8.3CH <i>ELQ8.3CH</i>	6QTH12
32	. PS_ELQ9.2CV <i>ELQ9.2CV</i>	6QTV12
BOP	50-20GL, Kepco	UQTVIZ
	50-20GL, Kepco . PS_ELD5.1-3 <i>ELD5.3</i>	6D20
1	50-20GL, Kepco PS_ELD5.1-3 ELD5.3 ELD5.2 ELD5.1	
1	50-20GL, Kepco . PS_ELD5.1-3 <i>ELD5.3</i> <i>ELD5.2</i> <i>ELD5.1</i> . PS_ELS16.1-4 <i>ELS16.1</i>	6D20 6D40 6D20 10S10
1	50-20GL, Kepco PS_ELD5.1-3 <i>ELD5.3</i> <i>ELD5.2</i> <i>ELD5.1</i> PS_ELS16.1-4 <i>ELS16.1</i> <i>ELS16.2</i> <i>ELS16.3</i>	6D20 6D40 6D20 10S10 10S10 10S10
1	50-20GL, Kepco PS_ELD5.1-3 <i>ELD5.2</i> <i>ELD5.1</i> PS_ELS16.1-4 <i>ELS16.1</i> <i>ELS16.2</i> <i>ELS16.3</i> <i>ELS16.4</i> PS_ELS4.1-2	6D20 6D40 6D20 10S10 10S10 10S10 10S10
1 2 3	50-20GL, Kepco PS_ELD5.1-3 <i>ELD5.2</i> <i>ELD5.1</i> PS_ELS16.1-4 <i>ELS16.1</i> <i>ELS16.2</i> <i>ELS16.3</i> <i>ELS16.4</i> PS_ELS4.1-2 <i>ELS4.2</i> <i>ELS4.1</i>	6D20 6D40 6D20 10S10 10S10 10S10
1 2 3	50-20GL, Kepco PS_ELD5.1-3 <i>ELD5.2</i> <i>ELD5.1</i> PS_ELS16.1-4 <i>ELS16.1</i> <i>ELS16.2</i> <i>ELS16.3</i> <i>ELS16.4</i> PS_ELS4.1-2 <i>ELS4.2</i>	6D20 6D40 6D20 10S10 10S10 10S10 10S10 10S10 11S10
1 2 3 4	50-20GL, Kepco PS_ELD5.1-3 <i>ELD5.3</i> <i>ELD5.2</i> <i>ELD5.1</i> PS_ELS16.1-4 <i>ELS16.2</i> <i>ELS16.3</i> <i>ELS16.4</i> PS_ELS4.1-2 <i>ELS4.1</i> PS_ELS4.3-4 <i>ELS4.3</i>	6D20 6D40 6D20 10S10 10S10 10S10 10S10 11S10 11S10
1 2 3 4	50-20GL, Kepco PS_ELD5.1-3 <i>ELD5.3</i> <i>ELD5.2</i> <i>ELD5.1</i> PS_ELS16.1-4 <i>ELS16.1</i> <i>ELS16.2</i> <i>ELS16.3</i> <i>ELS16.4</i> PS_ELS4.1-2 <i>ELS4.1</i> PS_ELS4.3-4 <i>ELS4.4</i>	6D20 6D40 6D20 10S10 10S10 10S10 10S10 11S10 11S10 11S10