

THE QUALITY FACTOR AND OTHER CONSIDERATIONS IN RADIATION PROTECTION AT THE AGS

S. V. Musolino

February 1989

Collider Accelerator Department
Brookhaven National Laboratory

U.S. Department of Energy

USDOE Office of Science (SC)

Notice: This technical note has been authored by employees of Brookhaven Science Associates, LLC under Contract No. DE-AC02-76CH00016 with the U.S. Department of Energy. The publisher by accepting the technical note for publication acknowledges that the United States Government retains a non-exclusive, paid-up, irrevocable, world-wide license to publish or reproduce the published form of this technical note, or allow others to do so, for United States Government purposes.

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or any third party's use or the results of such use of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof or its contractors or subcontractors. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

Accelerator Division
Alternating Gradient Synchrotron
BROOKHAVEN NATIONAL LABORATORY
Associated Universities, Inc.
Upton, NY 11973

Accelerator Division
Technical Note
AGS/AD/Tech. Note No. 316

THE QUALITY FACTOR AND OTHER CONSIDERATIONS
IN RADIATION PROTECTION AT THE AGS

S. Musolino

February 16, 1989

Introduction

The prompt radiation environment outside the primary and secondary AGS beams consists mainly of neutrons, muons, and electromagnetic components. The energy distribution and the relative numbers of each vary as a function of the beam intensity, energy, optics, angular position from the source, and shield density thickness.

Health Physics measurements performed throughout the experimental halls use certain simplifying assumptions to make conservative first approximations to the exposure conditions. In the majority of cases, the techniques used result in an overestimate of the dose to personnel. This paper will review how raw data are collected and what steps are taken to ensure that Health Physics data are conservative.

Instrumentation and Data Collection

During periods of extracted beam, two health physics instruments are used to measure pulsed (neutron) radiation fields. Fixed area monitoring is done with Fermilab Chipmunks and general hand held surveys are accomplished with the HPI-1010. Both instruments measure absorbed dose. The unit of absorbed dose is rads, where 1 rad = 100 ergs/g. To account for the difference in biological response between radiation species, the rad dose must be adjusted by the Quality Factor (Q) to convert the measurement into rem, the unit of dose equivalent, where:

$$H \text{ (rem)} = \text{rad} \times Q.$$

Even though the above relation looks simple, in practice it is non-trivial to evaluate precisely when instantaneous readout is required for routine radiation protection. Therefore, the method is simplified and designed to err conservatively.

As mentioned above, these devices respond in rad and it follows that it is not possible to readily separate the rads from neutrons and the rads from the gamma-muon response. Therein lies the problem. In addition, the Quality Factor for neutrons is energy dependent and varies between 2 and 10. Fortunately, we know from neutron spectroscopy done at the Cosmotron, AGS, CERN-PS, and other facilities, that the Quality Factor, on average at a high energy accelerator is about 5. So the rads measured by the chambers are assumed to be from only neutrons and are then multiplied by 5. The typical radiation fields around the AGS are about 50-50 neutron and gamma-muon. Muons deliver dose biologically the same as gamma or beta particles and have a Q of 1 since they also are minimum ionizing. Therefore, that portion of the dose equivalent is overestimated a factor of 5. In cases where the neutron Quality Factor is greater than 5, the inevitable gamma-muon component will compensate for the measurement error. Typically, the overestimate will be in the range of 50 - 100%.

Another factor that causes differences between the active instruments and badge results is the fact that the Chipmunks are often located at the "worst case" point of highest exposure and the occupants of a given area usually incur lower average dose rates than seen by the Chipmunk.

Conclusions

So when your film badge comes back low or minimal, or your local Chipmunk seems high, there really is not a wide variation. The difference resides mainly in the fact that the badge dose is realistic and Health Physics survey data are comfortably pessimistic. Of course, the survey data could be corrected on a daily basis, but for productivity reasons, we only convert it when detailed interpretation is required for planning purposes.