

AGS VACUUM SEALS

J. C. Schuchman

February 1967

Collider Accelerator Department
Brookhaven National Laboratory

U.S. Department of Energy

USDOE Office of Science (SC)

Notice: This technical note has been authored by employees of Brookhaven Science Associates, LLC under Contract No. AT-30-2-GEN-16 with the U.S. Department of Energy. The publisher by accepting the technical note for publication acknowledges that the United States Government retains a non-exclusive, paid-up, irrevocable, world-wide license to publish or reproduce the published form of this technical note, or allow others to do so, for United States Government purposes.

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or any third party's use or the results of such use of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof or its contractors or subcontractors. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

Accelerator Department
 BROOKHAVEN NATIONAL LABORATORY
 Associated Universities, Inc.
 Upton, L.I., N.Y.

AGS DIVISION TECHNICAL NOTE

No. 30

J.C. Schuchman
 February 2, 1967


AGS VACUUM SEALS

The following metal vacuum seals are presently installed in the AGS as shown below:


<u>Quantity & Description</u>	<u>Location</u>	<u>Mating Flanges</u>
(1) Lo-Load 8 1/2" I.D.	E-17 Upstream,	Standard Chamber to Port-Box
(1) Lo-Load "	E-16 "	Standard Chamber to 5ft Chamber
(1) Lo-Load 2 1/4" I.D.	E-17 Ion Gauge	Brass to S/S
(1) Lo-Load "	E-17 Moss Spec,	S/S to aluminum
(1) O-Ring 8 1/2" I.D.	E-12 Upstream,	Standard chamber to Bellows
(1) O-Ring "	E-19 "	" " "
* (1) O-Ring "	F-6 Downstream,	" " "
* (1) O-Ring "	F-9 Upstream,	" " "
(1) C-Ring 8 1/2" I.D.	E-13 Downstream,	Standard chamber to Port-Box
(1) C-Ring "	E-14 Upstream,	" " to 5' Chamber
(1) C-Ring "	E-15 "	" " to Bellows
(1) C-Ring "	E-15 Downstream,	" " to 5' Chamber
(1) C-Ring "	E-18 Upstream,	" " to "
(1) C-Ring "	E-20 "	" " to Bellows

* These two seals were installed on Nov. 30th, all others Feb. 1st.


Lo-Load Seal



O-Ring Seal



C-Ring Seal



For on the spot identification the edges of the flanges were marked with colored ink pencils to identify the type of metal gasket used.

Blue for Lo-Load seal

Green " O-Ring seal

Yellow " C-Ring seal

A more complete report of the AGS metal seal status and development will follow.

Distribution:

- V. Buchanan
- R. Dryden
- C. Gould
- J. Grisolia
- D. Hooper
- C. Lasky
- A. Maschke
- I. Polk
- A. van Steenberg