

BNL-104526-2014-TECH

AGS/AD/Tech Note No. 93;BNL-104526-2014-IR

CBM POLYETHYLENE FOIL CALIBRATION

L. N. Blumberg

April 1972

Collider Accelerator Department Brookhaven National Laboratory

U.S. Department of Energy

USDOE Office of Science (SC)

Notice: This technical note has been authored by employees of Brookhaven Science Associates, LLC under Contract No.AT(30-1)-16 with the U.S. Department of Energy. The publisher by accepting the technical note for publication acknowledges that the United States Government retains a non-exclusive, paid-up, irrevocable, worldwide license to publish or reproduce the published form of this technical note, or allow others to do so, for United States Government purposes.

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or any third party's use or the results of such use of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof or its contractors or subcontractors. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

Accelerator Department BROOKHAVEN NATIONAL LABORATORY Associated Universities, Inc. Upton, New York

AGS DIVISION TECHNICAL NOTE

<u>No. 93</u>

L. Blumberg, J. Gabusi, J. Glenn, G. Levine, D. Rahm May 9, 1972

CBM POLYETHYLENE FOIL CALIBRATION

On April 12, 1972 the full beam was extracted in one turn into the Slow External Beam channel. Polyethylene foils were exposed at F10 and A target station and a ratio of 1 CBM proton equaling .757 poly foil proton was measured.

The internal beam was measured by the CEM which had been calibrated on 1/28/72 at 1.69 volts out of the monitor per 1 volt p-p into the PUE cathode follower to give 1 volt per 10^{12} protons or 10^{10} protons per count on the CBM voltmeter.¹ There was no observable beam loss during the time of extraction when the fast kicker was not pulsed. The transmission efficiency of the bunches from the L10 fast kicker to the F7 electrode was 95% \pm 1% as measured from photographs of the F7 pick-up electrode signal before and after the kick. The radial position of the beam at F7 was 1.5 in. outside as measured by the F7 PUE and 2.2 in. outside as measured by a radioautograph of the F10 foil. With an outside aperture of 3 to 4 in. between F7 and F10 there should have been no beam aperturing. A spill counter at F10 calibrated by beam loss on the F10 flag indicated less than 0.1% loss in this area. Thus we assume no beam is lost from F7 to F10 and the transmission efficiency to F10 is 95% of the beam measured by the CBM.

A stack of three foils were exposed to the protons at both F10 and A on the first run and at A on the second run. The middle foil was counted. The A foils were measured and weighed, then constant area holes punched out and weighed. The F10 foil's "hot spot" was punched out and weighed. The density of foils was measured to better than 1%. The foils were counted in the chemistry well counter. The protons traversing the foils were calculated using

1. E. Gill, private communication (1972).

29.474 minutes for the mean life of 11 C, a gas loss correction 2 of .88, a counter efficiency of .74 and a cross section 3 of 25.9 x 10^{-27} cm 2 . The foils exposed at the A station have the equivalent of about 92 mils of aluminum upstream consisting of the A54 SEC, A91 split plates, a thin window, the A91 flag, and about 1 ft of air. From previous 4 tests of foil activation vs. thickness of aluminum absorber upstream of the foil, this material would make the A foil read 4.5% high. The foil exposures are tabulated below.

| <u>Run No.</u> | Device | Measured Protons <u>x</u> 10 ¹³ | $\frac{\text{Corrected Protons}}{\times 10^{13}}$ |
|----------------|----------|---|---|
| 1 | F10 foil | 1.267 | 1.267 |
| 1 | A foil | 1.327 | 1.270 |
| 2 | A foil | 1.185 | 1.134 |
| 2 | CBM | 1.577 | 1.498 |

From the first run the transmission efficiency from F10 to A is 100%. The second run gives a ratio of 1 CBM proton equaling $.757 \pm 1.5\%$ foil protons. This difference in "response" between the CBM and foils will be investigated further, using a calibrated current transformer as another method to measure proton intensity in the machine.

A recent (Feb. 1972) measurement of 60.6% for SEB efficiency (actually 58% when the build-up correction is applied) must now be corrected to 77%. The calculated⁵ theoretical SEB extraction efficiency with the present .030 in. F5 septum is 85.3%.

Distr: Department Administration AGS Division Staff EP&S Division Staff

-2-

^{2.} Cumming, et al. Phys. Rev. Lett 6, 484, (1961)

^{3.} J.B. Cumming, G. Friedlander and S. Katcoff, Phys. Rev. <u>125</u>, 2078, (1962).

^{4.} G.W. Bennett, private communication (1968).

^{5.} Blumberg, Barton, Bennett, Fox, Glenn, Hsieh, Nawrocky and Soukas, BNL Accel. Dept. Int. Rep. AGS DIV 69-8 (1969) and IEEE Trans. Nucl. Sci. <u>NS-16</u>, 234 (1969).