

## G-10 90ø MONITOR TELESCOPE UPDATE

J. Guthy

July 1972

Collider Accelerator Department  
**Brookhaven National Laboratory**

**U.S. Department of Energy**

USDOE Office of Science (SC)

Notice: This technical note has been authored by employees of Brookhaven Science Associates, LLC under Contract No.AT(30-1)-16 with the U.S. Department of Energy. The publisher by accepting the technical note for publication acknowledges that the United States Government retains a non-exclusive, paid-up, irrevocable, world-wide license to publish or reproduce the published form of this technical note, or allow others to do so, for United States Government purposes.

## **DISCLAIMER**

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or any third party's use or the results of such use of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof or its contractors or subcontractors. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

Accelerator Department  
BROOKHAVEN NATIONAL LABORATORY  
Associated Universities, Inc.  
Upton, New York

AGS DIVISION TECHNICAL NOTE

No. 97

J. Guthy, J. Glenn, G.W. Bennett  
July 26, 1972

G-10 90° MONITOR TELESCOPE

This note updates AGS Tech Note No. 89 dated January 14, 1972.

The G-10 monitor telescope has been replated and recalibrated. We now read 1878 counts per  $10^{12}$  protons. The high voltages are:

HVA = 2100

HVB = 1850

HVC = 2050

The change is partly due to recalibration of the circulating beam intensity monitor, aging and changes in temperature.

We expect to have this monitor recalibrated periodically.

Distr: Department Administration  
AGS Division Staff  
EP&S Division Staff  
AGS Experimenters

BROOKHAVEN NATIONAL LABORATORY

SHEET NO. 2 OF 2

JOB NO.

BY: *Anthony* DATE: *7/26/57*

CH'D. BY: *C* DATE: *7/26/57*

DEPT. OR PROJECT

