

CALIBRATION OF THE SECONDARY EMISSION CHAMBERS CE010 AND A300

V. Agoritsas

January 1976

Collider Accelerator Department
Brookhaven National Laboratory

U.S. Department of Energy

USDOE Office of Science (SC)

Notice: This technical note has been authored by employees of Brookhaven Science Associates, LLC under Contract No.E(30-1)-16 with the U.S. Department of Energy. The publisher by accepting the technical note for publication acknowledges that the United States Government retains a non-exclusive, paid-up, irrevocable, world-wide license to publish or reproduce the published form of this technical note, or allow others to do so, for United States Government purposes.

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or any third party's use or the results of such use of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof or its contractors or subcontractors. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

Accelerator Department
BROOKHAVEN NATIONAL LABORATORY
Associated Universities, Inc.
Upton, New York

AGS DIVISION TECHNICAL NOTE

No. 117

V. Agoritsas[†], J. Balsamo, R. Witkover

January 22, 1976


CALIBRATION OF THE SECONDARY EMISSION CHAMBERS
CE010 AND A300

I. Introduction

Since mid October 1975 a series of measurements have been made to intercalibrate beam intensity measuring devices and methods e.g. current transformers¹, secondary emission chambers^{2,3} and foil irradiation techniques⁴ in both fast ejected and slow ejected beams. A complete report of these measurements is in preparation. In this technical note, we report on the intercalibration of SEC CE010 and SEC A300 which was felt necessary after these two SEC's were interchanged during the January 1976 shutdown of the AGS.

II. Calibration of the SEC CE010, January 16, 1976


Three sandwiches of thin polyethylene and aluminum foils were prepared and installed inside the instrument box C12 in vacuum immediately downstream of the SEC CE010. Each sandwich consisted of 3 polyethylene and 3 aluminum foils. The fluxes of the 28 GeV protons were determined using the following two nuclear reactions:⁵



$$\left[\sigma = 2.5 \times 10^{-26} \text{ barns} \quad \text{and} \quad \lambda = \tau^{-1} (\ln 2)^{-1} = 3.3977 \times 10^{-2} \text{ min}^{-1} \right]$$

and

[†]Visiting Physicist from CERN.



$$\left[\sigma = 8 \times 10^{-27} \text{ barns} \quad \lambda = 7.7016 \times 10^{-4} \text{ min}^{-1} \right]$$

The activity of the foils was measured in the "W3" scintillation counter of the Chemistry Department which has an efficiency of 0.636 for ${}_{6}^{11}\text{C}$ radiation and 0.513 for ${}_{11}^{24}\text{Na}$ radiation. The SEC's were calibrated previously against the current transformer in the fast ejected proton beam and were set such that 1 SEC count = 10^9 protons \pm 5%.

During irradiation the SEC registered:

$$\text{SEC CE010} = 12219 \text{ counts}$$

which corresponds to a proton flux of

$$\text{UT799} = 1.222 \times 10^{13} \text{ protons.}$$

Proton Fluxes Derived From the Irradiated Foils

A) CH_2 (polyethylene) foils:

Sandwich A (normal to the beam)	foil #1	$I_p = 1.307 \times 10^{13}$	protons
	foil #2	$I_p = 1.290 \times 10^{13}$	
	foil #3	$I_p = 1.290 \times 10^{13}$	
		$\langle I_p \rangle = 1.295 \times 10^{13}$	
Sandwich B (45° to the beam, corr. for angle)	foil #1	$I_p = 1.277 \times 10^{13}$	
	foil #2	$I_p = 1.203 \times 10^{13}$	
	foil #3	$I_p = 1.308 \times 10^{13}$	
		$\langle I_p \rangle = 1.262 \times 10^{13}$	
Sandwich C (normal to the beam)	foil #1	$I_p = 1.438 \times 10^{13}$	
	foil #2	$I_p = 1.293 \times 10^{13}$	
	foil #3	$I_p = 1.314 \times 10^{13}$	
		$\langle I_p \rangle = 1.340 \times 10^{13}$	
All CH_2 foil mean		$\langle\langle I_p \rangle\rangle = 1.300 \times 10^{13}$	

and the ratio is: $\frac{\text{SEC CE010}}{\langle\langle I_p \rangle\rangle \text{CH}_2} = \frac{1.222}{1.300} = 0.94$

B) Aluminum foils.

Sandwich A (normal to the beam)	foil #1	$I_p = 1.275 \times 10^{13}$	protons
	foil #2	$I_p = 1.276 \times 10^{13}$	
	foil #3	$I_p = 1.276 \times 10^{13}$	
		$\langle I_p \rangle = 1.276 \times 10^{13}$	

Sandwich B (45° to the beam, corr. for angle)	foil #1	$I_p = 1.222 \times 10^{13}$
	foil #2	$I_p = 1.233 \times 10^{13}$
	foil #3	$I_p = 1.233 \times 10^{13}$
		$\langle I_p \rangle = 1.230 \times 10^{13}$

Sandwich C (normal to the beam)	foil #1	$I_p = 1.237 \times 10^{13}$
	foil #2	$I_p = 1.268 \times 10^{13}$
	foil #3	$I_p = 1.255 \times 10^{13}$
		$\langle I_p \rangle = 1.253 \times 10^{13}$

All Al foils mean $\langle\langle I_p \rangle\rangle = 1.250 \times 10^{13}$

All Al foils mean

and the ratio is: $\frac{SEC_{CE010}}{\langle\langle I_p \rangle\rangle_{Al}} = \frac{1.222}{1.250} = .97$

III. Calibration of the SEC in the "A Line" (A300)

Only one sandwich of 3 polyethylene foils and 3 aluminum foils were irradiated in the A line. The sandwich was placed several cm after the SEC in the A line.

During irradiation in SEC in "A line" registered:

$$SEC_{A300} = 10025 \text{ counts} = 1.0025 \times 10^{13} \text{ protons.}$$

The W3 scintillation counter was used for measuring both the CH₂ and Al foil activities. The CH₂ foil activities were measured by the authors (V. Agoritsas and J. Balsamo). The aluminum foil activities were measured by J. Burger. The results are:

A) Polyethylene foils	foil #1	$I_p = 1.016 \times 10^{13}$
	foil #2	$I_p = 1.001 \times 10^{13}$
	foil #3	$I_p = 1.023 \times 10^{13}$
		$\langle I_p \rangle = 1.013 \times 10^{13}$

$$\text{and } \frac{\text{SEC A300}}{\langle I_p \rangle} = \frac{1.0025}{1.013} = .99$$

B) Aluminum foils

$$\#1, 2, 3 \quad \langle I_p \rangle = 1.058 \times 10^{13} \text{ protons}$$

$$\frac{\text{SECA3000}}{\langle I_p \rangle} = \frac{1.0025}{1.058} = .95$$

Another sandwich of only three aluminum foils in the A line gave

$$\langle I_p \rangle = 5.239 \times 10^{13}$$

while the SEC A300 registered 50437 counts = 5.0437×10^{13} protons.

$$\frac{\text{SEC A300}}{\langle I_p \rangle} = \frac{5.0437 \times 10^{13}}{5.239 \times 10^{13}} = \underline{\underline{0.96}}$$

IV. Conclusion

The obtained results are within the relative expected discrepancies of the foil irradiation measurements. In our final report we shall discuss the observed discrepancies. Thus it must be concluded that to within the experimental variation, there is no difference in calibration between the calibration of the CE10 SEC and the A300 SEC, but the calibration may be as much as 5% low compared to foil measurements.

V. Acknowledgments

We are grateful to J.B. Cumming who permitted us to work day and night in his very well equipped counting room as well as for his valuable discussions. Thanks are due to Y.Y. Lee, A. Soukas and L. Repeta for their help.

References

1. G. Bennett and J. Guthy, "Current Transformers", FEB Operations Manual, page IV-A-1, March 1974.
2. V. Agoritsas and S. Battisti, "Operating Performance and Intercalibration of the Intensity Measuring Device of the CPS Circulating and Ejected Beams", CERN/MPS/CO 70-9, December 9, 1970.
3. R.L. Witkover, "New Instrumentation for the Slow Extracted Beams Users", BNL Accel. Dept. Int. Rep. AGS DIV 76-1, Feb. 10, 1976.
4. J.B. Cumming, "Monitor Reactions for High Energy Proton Beams", Ann. Rev. Nucl. Sci. 13, 261 (1963).
5. J.B. Cumming, private communication.

lsk

Distribution

Dept. Admin.
AGS Div. Physicists and Engineers
EP&S Div. Physicists
EP&S Div. Liaison Engineers & Experimenters