

AGS INTERNAL TARGETS

E. Gill

January 1978

Collider Accelerator Department
Brookhaven National Laboratory

U.S. Department of Energy

USDOE Office of Science (SC)

Notice: This technical note has been authored by employees of Brookhaven Science Associates, LLC under Contract No.EY-76-C-02-0016 with the U.S. Department of Energy. The publisher by accepting the technical note for publication acknowledges that the United States Government retains a non-exclusive, paid-up, irrevocable, world-wide license to publish or reproduce the published form of this technical note, or allow others to do so, for United States Government purposes.

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or any third party's use or the results of such use of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof or its contractors or subcontractors. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

Accelerator Department
BROOKHAVEN NATIONAL LABORATORY
Associated Universities, Inc.
Upton, New York 11973

AGS DIVISION TECHNICAL NOTE

No. 142

AGS INTERNAL TARGETS

E. Gill, S. Naase and J. Gabusi

January 24, 1978

Purpose

To run the internal targets at four locations.

The program is call TGT and located in R space R.

The data files are kept in 25, 27.

Locations

At present there are two horizontal locations, and one vertical. There are two targets at each location, In out, Top Bottom.

The horizontals are located at J-5 (Beta max) and D-15 (Beta min). The vertical target is located at J-19 (Beta max).

Calibration

The position is read in inches, on the target page where 0 is the center position and plus is outside for horizontal and up for vertical.

Operation

The drive is on the target page, CALL RTD15, RTJ05 and RTJ19.

To flip the target, it is necessary to run it from a terminal, and call in the function that is saved for that device. Once it is called and set up it can be switched on and off by target page. No other changes can be made from this page. The only other information on this page is the readbacks which tell the status of the target. The normal reading for a down target is 500 ± 200 counts. An up target 1500 ± 200 count.

A typical set up is shown on the printout, example A, which gives all the information on how to run and make changes. This set up shows how to flip the J50HI^{*}. The restore routine is used to go out and get the proper function from the data file and put it into either function A or B whichever one has been selected. This is always the first step in selecting a target. Once this is done, any of the routines can be changed by following the help call. The name of the data files are the same as the target page example C. To keep the target in the up position, first go out and get the data file for J50HI, then proceed to example B.

If you need different functions you can save them under different names in the data file and call back at will.

The hardware is arranged in such a way that no two targets on the same drive can be flipped together.

Example D shows a block diagram of the overall system, with drawing numbers for more detailed information.

Distr: Dept. Admin.
AGS Division S&P
MCR Operators
MCR Log

* J50HI (Location at J5 and Horizontal Inside)

R R

EXAMPLE A

#TGT


FUNCTION (A) : A

NAME (SPR1) : J5HI

RESTORE

H

<CAR.RET.>	LIST PRESENT SETUP
ON	TURN FUNC. GEN. ON
OFF	TURN FUNC. GEN. OFF
OUTPUT (OUT)	SEND FUNCTION TO GEN
ZERO	ZERO FUNCTION IN MEMORY
FUNCT (FG)	SELECT A FUNCTION GEN
NAME	SELECT A TARGET MOTOR
DEFINE (DEF)	SET TARGET MPX CHN#
SAVE	SAVE FUNCTION ON DISC
RESTORE	GET SAVED FUNC AND OUTP
UP	TGT STARTS UP NOW
DWELL	TGT RESTS FULLY ERECT
DELTA	VARY AMPLITUDES BY NZ
TIME	SET 6 FUNCTIONAL TIMES
AMP	SET 6 AMPLITUDES

^
Typical used routine.

Dwell

TYPICAL TARGET FLIPPING

Example 13

R R

To Keep target in up position

*TGT

FUNCTION (4) : A

NAME (SPR1) : J50HI

RESTORE

AMP

	<i>up</i>	<i>up brake</i>	<i>Hold</i>	<i>Down</i>	<i>Down brake</i>	<i>Down hold</i>
AMPLITUDES (1795,	-747,	448,	-1495,	298,	-372)
:	1000					

OUTPUT SENDS ONE VALUE to hold target up.

AMP

AMPLITUDES (1000,	0,	0,	0,	0,	0)
:	1000					

FINE

TIMES (100, 146, 157, 575, 655, 691)

EXAMPLE C

RIN	TARGET	SEL	24-JAN-78	11:12	43.1
EQUIPMENT	MODE	COMMAND	READBACK	AT	40
1.	1	D15HO	0	0	
1.	2	D15HI	0	0	
1.	3	J05HO	0	670	
1.	4	J05HI	0	1815	
1.	5	J19UU	0	580	
1.	6	J19UL	0	555	
1.	7	SPAT1	0	0	
1.	8	SPAT2	0	0	
1.	9	TFLPA	SPAT1	6	
1.	10	TFLPB	D15HO	0	
SEB	RTJ05	0	OFF	1869	ON
SEB	RTJ19	0	OFF	2242	ON

1 2 3 4 5 6 7 8
 SEL SET GETT SAUT READ BACK NEXT LIB

EXAMPLE D

BROOKHAVEN NATIONAL LABORATORY

BY E. GIL DATE 6/28
CHKD. BY DATE

SUBJECT BLANKS DISARM
INTERNAL TARGETS

SHEET NO. OF
JOB NO.

DEPT. OR PROJECT

RING

FLIP MOTOR

D11-PE2-3



DRIVE Motor

6.8A SURGE
2A D.C

10K 3T

Address 1367
1367
PPC

Select Motor



D11-E468-2

FUNCTION GEN. D09-E659-3

ADDRESS (FUNCTION GEN)
1376 PPC

ADDRESS
1377 PPC


D09-E834-3

READ TPT
CLT



MCB

Down = 50u ± 2ms
Up = 150u ± 2ms



Address 0055 PPB J5
0054 PPB J19
PPB D15