

THE AGS BPM SYSTEM UPGRADE

D. J. Ciardullo

January 1995

Collider Accelerator Department
Brookhaven National Laboratory

U.S. Department of Energy

USDOE Office of Science (SC)

Notice: This technical note has been authored by employees of Brookhaven Science Associates, LLC under Contract No. DE-AC02-76CH00016 with the U.S. Department of Energy. The publisher by accepting the technical note for publication acknowledges that the United States Government retains a non-exclusive, paid-up, irrevocable, world-wide license to publish or reproduce the published form of this technical note, or allow others to do so, for United States Government purposes.

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or any third party's use or the results of such use of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof or its contractors or subcontractors. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

Accelerator Division
Alternating Gradient Synchrotron Department
BROOKHAVEN NATIONAL LABORATORY
Upton, New York 11973

Accelerator Division
Technical Note

AGS/AD/Tech. Note No. 411

THE AGS BPM SYSTEM UPGRADE

D.J. Ciardullo and J.M. Brennan

January 31, 1995

The AGS BPM System Upgrade

Introduction:

It is desired to upgrade the existing AGS BPM system to accommodate the wide range of intensities accelerated within the AGS, including protons, heavy ions and polarized protons. The current goal is to design an average orbit system capable of operating with beam intensities from 10^9 to 10^{14} charges per pulse, using the presently installed AGS beam position detectors. To meet this challenge, new signal processing hardware needs to be designed. This document is a preliminary assessment of the basic objectives, philosophies and overall design concepts for a new average orbit system for the AGS. As such, feedback and discussion are encouraged.

System Objectives:

The following is a list of the key system objectives for this upgrade:

1. *Build an average orbit measurement system using the existing, recently modified PUE detectors.*
 - The PUE detectors currently used in the AGS for orbit acquisition have been upgraded in recent years. Signal wiring internal to the detectors have been replaced with radiation-hard shielded cable, and new vacuum flanges with BNC connectors have been installed. These detectors exhibit a linear differential response with beam position, and are considered to be of sufficient quality for use as the signal transducers for this upgrade.
2. *Measure average orbits for intensities from 10^9 to 10^{14} charges, over an RF range of 1.7MHz to 4.5 MHz, with PPM compatibility.*
 - This 5 decade span of intensities implies a basic dynamic range requirement of 100 dB; However, when the effects of non-zero beam position and bunch signal shape are taken into account, the dynamic range requirement for the peak voltage seen by a single PUE increases by an additional 20dB. [Note that beam positions several centimeters off-center can double the voltage seen on a particular electrode as compared with its "centered" value. In addition, the bunching factor for high intensity protons typically changes by a factor of five].
 - The system is required to hold its position measurement accuracy stable over greater than an octave frequency sweep, concurrently with the intensity range requirement.
 - PPM compatibility suggests that the system should not require any ring access for amplifier "change-overs" between runs.
3. *Build a reliable hardware system with a high degree of maintainability.*
 - This objective requires that there be no active electronics located inside the high radiation environment of the AGS ring, where high failure rates and lack of accessibility would dramatically increase system down time.

4. *Measure orbits accurately for a wide range of bunch lengths, from 10 nsec to 300 nsec.*
 - Since the existing PUEs are of the electrostatic variety, their signals are a representation of the longitudinal charge density in the beam bunch. As such, the peak amplitude of the PUE signals will vary over a wide range, effectively increasing the dynamic range required of the system.
5. *Measure average orbits with one or more bunches in the AGS (i.e., less than a "full" machine).*

Basic System Philosophies:

In an attempt to realize the system objectives, design and development of the AGS BPM system upgrade will adhere to the following basic system philosophies:

- **Keep active electronics out of the AGS ring!**
Radiation levels are too high for reliable operation. Also, the need for access makes repairs and/or "on-the-fly" modifications extremely difficult (proven with the Booster BPM system).
- **Build an *Average Orbit System*, not a universal PUE interface;**
This upgrade is not being designed to provide raw, wideband PUE signals. It will not provide trajectories (single turn orbits), and does not allow for bunch tracking. Due to the high dynamic range and broad bandwidth requirements of this upgrade, the purpose of the BPM system must be well defined in advance. The front end must essentially be a closed system in order to minimize corruption of the very low level PUE signals.
- **Design in a well defined partition between the instrumentation and data acquisition portions of the system;**
A well-defined division between these major sections increases the testability of each, as well as of the overall system. The processing electronics would provide as its output 216 channels of frequency-encoded position information (horizontal, vertical and "comfort" intensity). The Instrument Controller would then be responsible for inputting the FM information into the HP/Apollo system.
- **Use frequency-domain signal processing techniques;**
Position processing will be done in the frequency domain, operating on only one spectral component of the bunch signals from the AGS BPM detectors. With this method only a single component is needed, so there is no need to preserve the bunch shape fidelity of the PUE signals. This allows the use of heavy filtering techniques to boost their effective S/N ratio.
Narrowband rf processing also abolishes the need for fast, gated integrators; The implication here is that the bunch signals no longer need to be baseline restored. As such, the need for intricate, nanosecond timing is completely eliminated. In addition, the BPM measurement software does not have to compensate for the time dependant gain and offset errors normally associated with active integrators.

What the BPM Upgrade Would Include:

The AIP provides funding for an upgrade of the hardware between the PUE transformers and the Instrument Controller (inclusive). In order to better describe the proposed upgrade, the BPM system will be divided into three basic areas:

1. Design of a new PUE signal delivery system;
2. Design of new average orbit processing electronics;
3. Design of the centralized peripheral hardware used to route the real-time orbit information from all BPM locations into one location (e.g., the Instrument Controller, multiplexer, etc.).

Figure 1 shows a preliminary "system concept" diagram, indicating the location of the different sections of the upgrade hardware. A brief description of what would be included in the upgrade follows.

1. Design of a new PUE signal delivery system

To satisfy the contradicting requirements of processing low intensity bunch signals while keeping only passive electronics in the ring, a systems approach must be taken when delivering the bunch signals to the electronics front end. Of paramount importance is the use of high quality (i.e., low loss, solid shield) coaxial cable; The PUE signals, as well as the electronics which process them, will be completely enveloped by an electrostatic enclosure (what can be thought of as a "faraday cage"). The plan is to use solid shield coaxial cable (which provides 100% rf shield coverage) to bring the PUE signals into the A-10, E-10 and H-10 houses for analog processing. The processing hardware for each set of four BPM signals will utilize extruded aluminum-type enclosures in such a manner as to approximate, as closely as possible, one continuous extended electrostatic shield. This effectively prevents any rf traveling up the shield of the coax from interfering with the processing electronics.

The upstream end of the signal delivery system includes new 16:1 wideband transformers. Since they are part of the PUE signal delivery system, the input and output side of these transformers are purposely *not* isolated. The function of these devices is to broadband match the impedance of each PUE to a resistive 50Ω (enabling the use of commercially available 50Ω cable), while minimizing the amount of signal loss due to capacitive electrode loading. At high frequencies, the transformer assemblies will terminate the 3nsec coaxial lead-out cables and suppress the quarter-wavelength resonance at $\approx 85\text{MHz}$. A preliminary design for an appropriate wideband autotransformer is currently being tested in the AGS.

2. Development of new average orbit processing electronics

It is planned to distribute the processing electronics equally among the A-10, E-10 and H-10 houses. This satisfies the requirement of bringing the active components of the system out of the ring, while at the same time minimizing the lengths of the signal cables. For purposes of economy, cable runs from different detector locations will not be of matched length; However, the four PUE cables from each detector location need to be phase matched to one another.

The processing electronics located in the X-10 houses would provide localized gain and

average orbit processing for 72 BPM detectors, each with dual-plane measurement capability. Current development efforts are directed toward designing a very high dynamic range rf receiver-based processing unit with continuous, automatic gain adjustment. In addition to an uninterrupted flow of average orbit information, this architecture also allows for a "comfort" output which indicates the beam intensity (This output is basically the AGC error voltage). Observation of this output gives an indication of whether or not the signal level is sufficient for successful operation of the system. All three analog outputs (horizontal position, vertical position and log intensity) are isolated using fiberoptic links with digital frequency modulation. Additional benefits of using this technique include a very high degree of noise immunity and protection against the amplitude variations normally associated with fiberoptic linearity and attenuation.

Localized test modules will be used for purposes of system test. This BITE (Built In Test Equipment) circuitry functions to verify proper operation of the remote processing electronics. RF calibration and Local Oscillator (LO) signals are normally provided by the existing AGS LLRF system, and are shipped to the X-10 houses via analog fiberoptics. [A backup swept signal/LO pair would also reside in building 929, for the purpose of providing simple "go/no go" tests when the LLRF system is off-line]. Attenuation of the individual rf test signals would occur at each set of processing electronics, so that variations in both intensity and position could be simulated. The BITE circuitry would provide:

1. Routing control of the switches which deliver the calibration signal to the processing electronics;
2. Swept-frequency rf calibration signals and LO (for use when the LLRF system is off-line);
3. Common-mode amplitude control (in discrete steps) of the test signals to verify system dynamic range;
4. Differential-mode amplitude control of the test signals to simulate several different beam positions;
5. LO distribution, and distribution of the rf and LO test signals.

3. Design of centralized peripheral hardware

The peripheral hardware for the upgrade would reside in building 929 (possibly above the mezzanine). Plans include the design of a digital multiplexer to provide the MCR with 4 FM signals of any combination of horizontal, vertical or intensity signals. Sending the FM signals to the MCR helps preserve signal integrity, but will require a piece of equipment at the receiving end to convert the modulated signals back to analog. The mux could also be used as a convenient "tie point" where the orbit information from any one or more BPM locations could be combined to obtain a true average orbit for the radial control loop. [This would decrease the sensitivity of the radial loop to loss of any one particular BPM location, but implies a position measurement bandwidth of 20kHz].

An Instrument Controller (also located in building 929) is required to sample and input to the HP/Apollo system the real-time orbit information from all BPM locations. Since the orbit information is always available, the sample times will be specified as time line events by the high level software. As such, the controller would include the hardware necessary to decode timeline events. The controller will receive frequency modulated orbit data directly, then use timed counters to evaluate the data; This eliminates the need to convert the information back to analog first, which tends to introduce additional errors into the data. Information from a total of 216 channels (72 locations x 3 channels/location) would need to be gathered by the controller.

In addition to gathering data from the BPM system, the Instrument Controller will control the multiplexer described above. Hardware commands are also needed to set-up the BITE functions for system test. The following is a preliminary list of required controller commands/readbacks [all signals are active low]:

<u>Command Name</u>	<u># Bits</u>	<u>Signal Type</u>	<u>Function</u>
/TEST_MODE	1	open collector	Puts electronics into testing mode.
/SIM_POS	4	open collector	Sets the diff mode gain of the test signals to simulate one of 16 discrete positions.
/SIM_POS_STAT	4	open collector	Status readback for SIM_POS command
/SIM_INT	4	open collector	Sets the common-mode gain of the test signals to simulate one of 16 discrete intensities.
/SIM_INT_STAT	4	open collector	Status readback for SIM_INT command
/CAL	1	O.C., 1 μ s pulse	Informs system to output a zero-position test signal 5 milliseconds before T0 .
Mux bits	32	open collector	This is a first-pass guess based on an 8 bit word (to select one of 216 lines) for each of 4 lines to the MCR. May be refined further at a later time.
Spare bits	8	open collector	Spare bits for possible later use
Spare pulse	4	O.C., 1 μ s pulse	Spare pulses for possible later use

AGS BPM UPGRADE OVERALL BLOCK DIAGRAM




Figure 1