

BNL-99235-2013-TECH C-A/AP/82;BNL-99235-2013-IR

Analysis and Measurement of Cross-Talk in a Superconducting cavity

Y. Zhao

August 2002

Collider Accelerator Department Brookhaven National Laboratory

U.S. Department of Energy

USDOE Office of Science (SC)

Notice: This technical note has been authored by employees of Brookhaven Science Associates, LLC under Contract No.DE-AC02-98CH10886 with the U.S. Department of Energy. The publisher by accepting the technical note for publication acknowledges that the United States Government retains a non-exclusive, paid-up, irrevocable, world-wide license to publish or reproduce the published form of this technical note, or allow others to do so, for United States Government purposes.

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or any third party's use or the results of such use of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof or its contractors or subcontractors. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

C-A/AP/#82 September 2002

ANALYSIS AND MEASUREMENT OF CROSS-TALK IN A SUPERCONDUCTING CAVITY

Yongxiang Zhao



Collider-Accelerator Department Brookhaven National Laboratory Upton, NY 11973

ANALYSIS AND MEASUREMENT OF CROSS-TALK IN A SUPERCONDUCTING CAVITY

Yongxiang Zhao

Physics Department, Bldg. 901-A Brookhaven National Laboratory, Upton, NY 11973

Abstract. A superconducting cavity used in a microwave gun requires that the launcher and the pickup probes be on the same side of the cavity, which causes direct coupling between them, or crosstalk. At room temperature, the crosstalk causes serious distortion of the RF response. This note addresses the phenomenon, the simulation results and the analysis, so that one can extract the desired information from the confusing signal.

INTRODUCTION

In the "Electron Cooling" project, a microwave gun is employed as an electron source, and a superconducting cavity (SCC) is used in the first experiment. Microwave guns have been successful in many places, such as at the ATF at BNL. Superconducting cavities also have been successfully employed in many laboratories around the world. However, to our knowledge, there is little experience of a SCC used in a microwave gun.

A SCC in an accelerator usually has its input antenna (or launcher) on one side and the pickup probe on the other side. The coupling between them is only through the cavity cells, and has a perceivable coupling only near the resonant frequency. For a cavity of a microwave gun, one end must be the cathode; thus, both launcher and pickup must be on the same side.

A schematic of the SCC is shown in Fig.1. Probe A is the input or launcher; probe B is the pickup; and probe C is an antenna used for temporary measurement only. For convenience of machining and welding, the launcher and pickup are diametrically opposite. The direct coupling generating a crosstalk becomes inevitable. It becomes more significant because of the fact that for a SCC the coupling between launcher and the cavity must be very weak at room temperature, and the external Q must be very large in order to match the cavity when it is cooled down to superconducting status. When the cavity is superconducting, crosstalk may be negligible small, but we have to test all the parameters of the cavity at room temperature, where the crosstalk is troublesome.

Our first measurements found that the pickup signal had too much distortion from a resonant curve. Fig. 2 shows some typical response of S21. We determined that the abnormal phenomenon was due to crosstalk. To extract the useful information from the undesirable signal is a challenge.

To this end, we had to make a model for simulation. Besides, analyses are also necessary in order to understand the relationship between parameters and the responses in the measurements. The formulae for calculation are collected. Finally, test results of the mentioned cavity are presented.



Fig. 1 Test setup

PHENOMENA IN THE MEASUREMENT

For a typical cavity with one port, the coupling and the Q value can be measured by the VSWR on the input line (or waveguide). However, the coupling in our case was too weak, so that the great majority of power reflects from the cavity and the conventional method results in too much error.

In the beginning, we measured the S21 of the cavity by means of the network analyzer HP8753D, with ports 1 and 2 connected to the launcher and pickup (probes A and B in Fig.1), respectively. Fig. 2(a) shows the magnitude of S21 (in dB). The peak frequency is about 1.3 GHz as expected, but the shape is far from the expected resonant curve. There is a dip frequency a little lower than the peak frequency. Near the peak the curve is markedly asymmetric so much so that the Q value calculated by the network analyzer does not make sense. Fig. 2(b) shows its phase response. Instead of an "S" curve with 180 degree span as it should be, it has much less span and looks like a "stroke curve", making things more confusing.

The pickup probe was a loop, which usually is regarded as a magnetic coupler. In theory it should not couple with the launcher, which is an electric coupler. Unfortunately, this was not true. When we rotated the loop, the responses of S21 were different in all four quadrants. At one position, the dip frequency moved to the upper side, as shown in Fig. 2(c); meanwhile, the phase response was also reversed, as shown in Fig. 2(d).

The fact that the pickup signal depended on the rotation angle of the loop implies that the loop involves both magnetic and electric couplings, just like a loop directional coupler. This is understandable, because the loop is located in a region of electric field where the magnetic field is very weak.

We found the crosstalk in this structure to be inevitable. Fortunately, we recognize that this crosstalk will diminish when the cavity is cooled down to its superconducting state (See later for detail.). We then replace the loop pickup by another probe for easier machining and control.



Fig.2. Some typical measured S21 responses of the cavity with different crosstalk. The left column displays the magnitudes of S21; and the right column, their phases.

In order to make a better measurement at room temperature, we employed a third probe (probe C in Fig. 1) made of a semi-rigid coaxial cable, because it can be inserted close to the cavity to increase the coupling with less crosstalk. Fig. 2(e) and (f) show the results by using probe C and probe A. The dip frequency shifts away from the peak that the resonant curve becomes meaningful.

The following simulation and analysis explain these phenomena.

MODELING AND SIMULATION

SUPERFISH has been used to simulate the cavity. ^[1] However, it cannot simulate coupling, which is a 3-dimensional problem. One can employ a 3-D code such as MAFIA if the field analysis is of concern.

However, in the present case, the equivalent circuit model is more convenient and can be simulated by PSpice. Especially, it gives a clear physical concept to help direct the design. The cavity can be represented by an L-C-R resonant circuit. Its equivalent parameters can be deduced from the output of SUPERFISH as:

 $C_0 = 1.048 \text{ pF}$ $L_0 = 14.30 \text{ nH}$ $R_0 = 9.34e5 \text{ ohm (room temp)}$ R/Q = 116.8 ohm



Fig. 3. Equivalent circuit of the cavity and couplers (a), with crosstalk added (b)

Fig. 3(a) shows the equivalent circuit, where C_0 - L_0 - R_0 represents the cavity. C1 and C2 represent the input and pickup couplers, respectively, with each connected to a 50 ohm cable. The values are arbitrarily defined for the moment. A capacitance C12 is added in Fig. 3(b). It represents a direct coupling between input and pickup and introduces crosstalk.

Here all the couplings are attributed to capacitance, because the probes are of rod shape and located in the electric field area. This may not always be true, as mentioned above, if the pickup is a loop. The latter case is complicated and will be dealt with in the next section.

Fig. 4 shows the results from PSpice. Evidently, without direct coupling, the curves

display normal resonance. The phase changes 180 degrees when the frequency crosses the resonance. When C12 is added (C12 = 0.001pF is chosen in this curve), both the magnitude and phase display are distorted, exactly as we measured.



Fig. 4. The simulation results of circuits Fig. 3(a)--green, and Fig.3(b)--red.

The dip frequency is due to the interference of two signals. One is the normal coupling through the cavity, of which the phase is very frequency sensitive. The other is due to the direct coupling, which is not sensitive to frequency. When the two signals had opposite phases, the signals cancel each other and thus formed a dip.

If one increases the Q of the cavity by cooling it down from room temperature to its superconducting state, in the vicinity of the resonant frequency the coupled signal through the cavity becomes much stronger than that through C12, and the crosstalk becomes less important. The response will revert to a normal resonance.



Fig. 5. The same direct coupling but different cavity Q

Fig. 5 shows the simulated circuit with (a) $R_0 = 900k\Omega$ and (b) $R_0 = 2000M\Omega$, representing room temperature and superconducting respectively. The PSpice output is shown in Fig. 6: the peak is much sharper. Meanwhile, the phase changes 180 degree



when the frequency crosses the resonance, although the direct coupling (C12) remains the same.

Fig. 6. S21 of different Q value, results of Fig. 5(a)--red and Fig.5(b)--green.

This verifies the above argument that the crosstalk becomes less important when the cavity is superconducting.

To check the effect of the direct coupling, Fig. 7 shows the response with different values of C12.



(a) S21 magnitude

(b) S21 phase

Fig. 7. S21 response with different values of C12: (1 -- 0.005pF, 2 -- 0.002pF, 3 -- 0.001pF, 4 -- 0.0002pF)

It shows that the weaker the direct coupling (the smaller C12), the larger the frequency separation between peak and dip, also the larger the phase expansion. When it is weak enough, as in curve 4, the dip frequency is too far away to be evident in the figure, meaning that the crosstalk is negligible. Meanwhile, the phase expanded to 180

degrees.

If the direct coupling is very strong, then not only will the resonant curve be distorted, but also the magnitude of S21 will increase and the peak frequency shift to a higher one.

ANALYSIS

The simulation gives a clear response, but doesn't give an explicit relationship between its parameters. Therefore, it is necessary to analyze the equivalent circuit. Fig. 8 shows the model in question. Fig. 9 gives a generalized form, which resembles a bridge circuit.





Fig. 9. The generalization of Fig.8

Applying Kirchhoff law, after algebraic manipulation one obtains the following matrix equation.

$$\begin{pmatrix} Y_1 & -B_{12} & -B_1 \\ -B_{12} & Y_2 & -B_2 \\ -B_1 & -B_2 & Y_0 \end{pmatrix} \begin{pmatrix} V_1 \\ V_2 \\ V_C \end{pmatrix} = \begin{pmatrix} G_s \\ 0 \\ 0 \end{pmatrix} V_g , \qquad (1)$$

where
$$Y_1 = G_S + B_1 + B_{12}$$

 $Y_2 = G_L + B_2 + B_{12}$
 $Y_0 = Y_C + B_1 + B_2$. (2)

 Y_0 is the total admittance of the cavity including the coupling capacitance. In our special case, $G_S = G_L = G_0 = 1/50$ (ohm), and the couplings are very weak such that B_1 , B_2 , B_{12} are negligible in comparison with G_0 . Then approximately,

$$Y_1 = Y_2 = G_0 = 0.02 \text{ mho}$$
(3)

Solving the equation (1), one obtains:

$$\frac{V_2}{V_g} = \frac{Y_0 B_{12} + B_1 B_2}{G_0 Y_0 - B_1^2 - B_2^2} \,. \tag{4}$$

The coupling coefficient β related to the coupling capacitance can be deduced by the ratio of power loss inside and outside the cavity in the absence of crosstalk. From the

parameters in Fig. 8 one obtains the following formulas.

$$\beta_{2} = \frac{P_{L}}{P_{C}} = \frac{V_{C}^{2} (R_{2}\omega_{0}C_{2})^{2}/2R_{2}}{V_{C}^{2}/2R_{C}} = R_{2}R_{C}(\omega_{0}C_{2})^{2}$$

$$= \frac{R_{C}R_{2}}{L_{C}C_{C}}C_{2}^{2} = Q_{0}^{2}\frac{R_{2}}{R_{C}}\frac{C_{2}^{2}}{C_{C}^{2}}$$
(5a)

Similarly

 $\beta_1 = Q_0^2 \frac{R_s}{R_c} \frac{C_1^2}{C_c^2} \,. \tag{5b}$

Note that $R_S = R_2 = 1/G_0$, and $R_c \omega_0 C_c = Q_0$.

Now we can explore the properties of formula (4). Only Y_0 is frequency sensitive. The other parameters can be regarded as approximately constant. In general, a ratio of polynomials can be characterized by its poles and zeros. Without crosstalk (i.e. $B_{12} = 0$), the numerator in formula (4) becomes constant and thus has no zero, corresponding to a simple resonance.

With crosstalk one needs to substitute the parameters in Fig.8:

$$B_1 = j\omega C_1, \quad B_2 = j\omega C_2, \quad B_{12} = j\omega C_{12}.$$
 (6)

After tedious algebraic manipulation that formula (4) can be rewritten in the form:

$$\frac{V_2}{V_g} = \frac{j\omega C_{12}}{G_0(1+\beta)} \cdot \frac{1+jQ_z \cdot 2(f-f_z)/f_0}{1+jQ_L \cdot 2(f-f_0)/f_0},$$
(7a)

where

$$\beta = \beta_1 + \beta_2$$
 ($\beta = \beta_1 + \beta_2 + \beta_3$ if there are 3 probes) (7b)

$$Q_L = \frac{Q_0}{1+\beta},\tag{7c}$$

$$Q_Z = Q_0 \left(1 + \delta_Z \right), \tag{7d}$$

$$f_Z = f_0(1 - \frac{1}{2}\delta_Z),$$
 (7e)

and

$$\delta_{Z} = \frac{C_{1}C_{2}}{C_{0}C_{12}}.$$
(7f)

There is a pole at $f = f_0$, and a zero at $f = f_Z$. This verifies what we have observed in the measurement. Around the resonant frequency f_0 , the Q is Q_L . Around f_Z , the dip shows an anti-resonance with Q of Q_Z . However, if the crosstalk is serious, f_0 is close to f_Z , and the resonant curve is so distorted that the peak frequency f_{max} is not exactly at f_0 , f_{min} is not exactly at f_Z , and the measured Q from 3 dB bandwidth is not exact Q_L .

Here we have introduced the important parameter δ_Z , relating the relative frequency separation of f_Z and f_0 . (see (7e))

$$\delta_{Z} = \frac{2(f_{0} - f_{Z})}{f_{0}}$$
(8)

The closer the two frequencies, the more the distortion of the resonant curve, implying more serious crosstalk.

In order to give a clear physical meaning to δ_Z , let's rewrite (7f) as:

$$\delta_{Z} = \frac{(C_{1}/C_{0})(C_{2}/C_{0})}{(C_{12}/C_{0})}$$
(9)

The numerator is a measure of the coupling through the cavity, while the denominator is a measure of the direct coupling. Thus the ratio gives a measure of the interference. The larger C12, the smaller δ_Z , and the stronger the crosstalk.

An implicit fact for forms (7) and (8) is that all the couplings are capacitive, as displayed by Fig.(8). Consequently, δ_Z is positive. However, it is not required by formula (4).

If the pickup coupling is primarily magnetic, then $B_2 = j\omega C_2$ in (6) will be replaced by $B_2 = -j\omega C_2 = -j/(\omega L)$; namely the sign is changed. (The symbol C or L is not important, because it's only a constant in the vicinity of the resonant frequency; only the sign is important.) In this case δ_Z is negative. Then from (7e), $f_Z > f_0$; i.e. the dip frequency is higher than the peak frequency. Meanwhile, the phase of S21 changes sign too (See later for detail.). This explains the phenomena observed in Figs. 2(c) and (d).

In addition, from formulas (7a), (7c) and (7d) one can see that $Q_L < Q_0$, while $Q_Z > Q_0$. This implies that the dip (or anti-resonance) is sharper than the peak, agreeing with what we've observed.

The agreement with observation once again verifies the circuit model. In order to get more general and convenient formulas, we can make a further normalization:

$$\delta = \frac{2(f - f_0)}{f_0},$$
 (10a)

$$F = Q_0 \,\delta,\tag{10b}$$

and

and

$$F_Z = Q_0 \,\delta_Z \,. \tag{10c}$$

Each has a physical meaning. δ is the relative frequency deviation. F is the normalized frequency deviation, such that F = 1 corresponds to a frequency at the edge of the 3-dB bandwidth. F_Z is the normalized crosstalk parameter that measures the deviation of the dip frequency from the resonant frequency, It gives a good measure of the crosstalk.

Substituting (5), (7f) and (10), then (7a) can be rewritten as:

$$\frac{V_2}{V_g} = \frac{j\sqrt{\beta_1\beta_2}}{1+\beta} \cdot \frac{1}{1+jk_LF} \cdot \frac{1+jk_Z(F+F_z)}{F_Z}$$
(11)

where k_Z and k_L are constants very close to unity:

$$k_Z = l + \delta_Z \approx 1 \,, \tag{12a}$$

$$k_L = 1/(1+\beta) \approx 1. \tag{12b}$$

Equation (11) has a clear physical meaning. On the RHS, the first two fractions represent the response in the absence of direct coupling. The first one is the magnitude of the coupling at resonant frequency. The second manifests the resonant performance. The third fraction involving F_Z manifests the effect of the direct coupling or the crosstalk. It approaches unity for $F_Z \rightarrow \infty$, i.e. whenever either Q_0 or δ_Z is very large. This verifies that F_Z is an even better measure of crosstalk than the parameter δ_Z .

At room temperature, F_Z is in a range that the crosstalk must be taken into account. When cooled down to the superconducting state, Q_0 increases a few orders of magnitude, such that the condition $F_Z >> 1$ is realized, and the crosstalk vanishes.

THE RELATIONSHIP TO THE TESTED PARAMETERS

Using the analyses above, we now are ready to calculate the parameters we want from the parameters that we can measure.

The parameters we want to know are the resonant frequency f_0 , Q_0 , coupling coefficient β , and external Q. Without crosstalk, for a simple resonator, they can easily be determined by the network analyzer. But, for the case concerned, neither can be measured directly without significant error.

The measurable parameters are S11 and S21, with corresponding frequencies. (see Fig.12).

Let's recall briefly the measurement of the coupling coefficient β of a normal cavity. It's simply determined by the SWR at the input transmission line; it's a function of frequency. (Using symbol ρ = SWR for short,.)

or
$$\beta = \rho(f_0)$$
 (over-coupling),
 $\beta = 1/\rho(f_0)$ (under-coupling). (13)

In our case, coupling is designed to be close to critical ($\beta = 1$) in the when in superconducting state, while at room temperature, β is of the order of 10^{-4} . In this case, the SWR is of the order of 10^{4} , which is too high to measure accurately. The SWR looks like Fig. 10. The SWR at off-resonance is only about 17, though in principle, for a lossless transmission line, the SWR is infinite.

The finite SWR implies the existence of a residual loss somewhere in the transmission line. Assume an equivalent lumped resistance R_r at the entrance of the cavity, as shown in Fig. 11. At off-resonance there is no power entering the cavity, but it is not a perfect open circuit due to residual losses. Then the measured SWR is

$$\rho_l = R_r. \tag{14a}$$

At resonance the cavity will introduce a minor extra loss, whereupon the SWR is

$$\rho_2 = \frac{R_r R_c}{R_r + R_c} \tag{14b}$$

One can easily find the β , which equals the SWR due to the cavity only:

$$\beta = \frac{1}{R_c} = \frac{\rho_1 - \rho_2}{\rho_1 \rho_2}$$
(15)

This formula gives a rough estimation of very weak coupling. However, the precision is not satisfactory, due to the error of the measured ρ s.



Fig.12 The measurable parameters

We should rely mostly on the measurement of S21.

Note that Vg in the equivalent circuit Fig.9 is twice the injection voltage on the input transmission line, or $2V_+$; the scattering parameter S21 is

$$S21 = \frac{V_2}{V_+} = \frac{2V_2}{V_g}.$$
 (16)

In a network analyzer, S21 usually reads in dB for magnitude and phase. The magnitude is:

$$S21 = 20\log \left| \frac{V_2}{V_+} \right| = 20\log \left| \frac{2V_2}{V_g} \right|$$
 (dB). (17)

From (11) we obtain:

$$S21 = 10\log\left\{\frac{4\beta_1\beta_2}{1+\beta_1+\beta_2} \cdot \frac{1}{1+k_L^2F^2} \cdot \frac{1+k_z^2(F+F_z)^2}{F_z^2}\right\}.$$
 (18)

The phase is:

$$\Phi(S21) = \arctan[k_z(F + F_z)] - \arctan(k_L F).$$
(19)

Note that both k_L and k_Z are very close to unity and thus negligible in most cases. $\beta = \beta_I + \beta_2 \ll 1$, also is negligible in many cases. Obviously, if crosstalk is negligible (i.e., $F_z \gg 1$), the maximum S21 occurs at resonance or F = 0, one obtains

$$S21_{max} = 10 \log (4\beta_I \beta_2).$$
 (20)

This well-known formula applies in the absence of crosstalk.

However, if the crosstalk is not negligible, the last two fractions in (18) must be taken into account. The maximum S21 does not occur at resonant frequency. Differentiating (18) with respect to *F*, one can find the peak and dip frequencies to be:

$$F_{\rm max} = \frac{\sqrt{F_z^2 + 4} - F_z}{2} , \qquad (21a)$$

$$F_{\min} = -\frac{\sqrt{F_z^2 + 4 + F_z}}{2},$$
(21b)

where F_{max} and F_{min} are normalized frequencies as defined in (10), or

 $F_{min} = Q_0 \cdot 2(f_{min} - f_0) / f_0$.

$$F_{max} = Q_0 \cdot 2(f_{max} - f_0) / f_0, \qquad (22a)$$

(22b)

and

Note that
$$F_{max}$$
 is always positive, meaning the peak frequency is not equal to the resonant frequency f_0 (pole point) except when crosstalk is negligible ($F_Z >> 1$). This deviation is

$$f_{\max} - f_0 = \frac{f_0}{2Q_0} \cdot \frac{\sqrt{F_z^2 + 4} - F_z}{2}.$$
 (23)

This deviation was found to be up to 70 kHz in some cases. Similarly, the dip frequency f_{min} also deviates from the "zero point" frequency f_z .

As seen in Fig. 12, f_{max} and f_{min} are measurable. From (22) and (21) their difference can be used to determine the crosstalk parameter F_Z , providing Q_0 is known.

$$F'_{z} = F_{\max} - F_{\min} = Q_0 \cdot 2(f_{\max} - f_{\min}) / f_0,$$
 (24a)

$$F_Z = \sqrt{F_z^{\,2} - 4}$$
. (24b)

.

and

Substituting (21) into (18), $S21_{max}$ and $S21_{min}$ are found to be:

$$S21_{\max} = 10 \cdot \log \left(\frac{4\beta_1 \beta_2}{1+\beta} \cdot \frac{1}{F_z^2} \cdot \frac{\sqrt{F_z^2 + 4} + F_z}{\sqrt{F_z^2 + 4} - F_z} \right),$$
(25a)

and

$$S21_{\min} = 10 \cdot \log\left(\frac{4\beta_1\beta_2}{1+\beta} \cdot \frac{1}{F_Z^2} \cdot \frac{\sqrt{F_Z^2 + 4} - F_Z}{\sqrt{F_Z^2 + 4} + F_Z}\right).$$
 (25b)

In practice, the signal S21_{min} is usually too weak to be reliable. Instead, one can measure a middle value S21_M at the middle frequency f_M , i.e.

$$f_M = \frac{f_{\text{max}} + f_{\text{min}}}{2}, \qquad F_M = \frac{F_{\text{max}} + F_{\text{min}}}{2} = \frac{F_Z}{2}$$
 (26)

and

$$S21_{M} = 10 \cdot \log\left(\frac{4\beta_{1}\beta_{2}}{1+\beta} \cdot \frac{1}{F_{Z}^{2}}\right)$$
(27)

Then

$$\Delta S_{0M} = S21_{\text{max}} - S21_M = 10 \cdot \log \frac{\sqrt{F_Z^2 + 4 + F_Z}}{\sqrt{F_Z^2 + 4 - F_Z}}$$
(28)

Defining $C = 10^{\Delta S_{0M}/20}$, (29a)

 ΔS_{0M} is a measured quantity, so *C* is a measurable parameter related to crosstalk From (28) we have

$$C = \frac{\sqrt{F_Z^2 + 4} + F_Z}{2}$$
, and $C^2 = \frac{\sqrt{F_Z^2 + 4} + F_Z}{\sqrt{F_Z^2 + 4} - F_Z}$. (29b)

Solving the equation, one can find F_Z from the measurable parameter ΔS_{0M} and C:

$$F_Z = C - \frac{1}{C} \tag{30}$$

According to (25a), (27) and (28), the approximate (20) should be corrected as

$$S21_{max} = 10 \log (4\beta_1 \beta_2) + \Delta S_0, \tag{31}$$

$$\Delta S_{0} = 10 \cdot \log \left(\frac{1}{F_{Z}^{2}} \cdot \frac{\sqrt{F_{Z}^{2} + 4} + F_{Z}}{\sqrt{F_{Z}^{2} + 4} - F_{Z}} \right) = 10 \cdot \log \frac{C^{2}}{F_{Z}^{2}}$$

$$= -20 \cdot \log \left(1 - \frac{1}{C^{2}} \right)$$
(32)

where

From S21_{max} one can easily find the product $\beta_1\beta_2$. We have to employ three probes, as shown in Fig.1. We can make different combinations, such that two of them connect to the two ports of the network analyzer, and measure the S21 for each combination. Once one measures the three products $\beta_A\beta_B$, $\beta_A\beta_C$, and $\beta_C\beta_B$, it is easy to find each individual

 β value by

$$\beta_{A} = \sqrt{\frac{(\beta_{A}\beta_{B}) \cdot (\beta_{A}\beta_{C})}{(\beta_{C}\beta_{B})}}$$
(33)

As mentioned before, the phase of S21 displays a stroke when crosstalk is not negligible. The magnitude of the phase stroke can be found by seeking the maximum of the $\Phi(S21)$ from (19). Differentiating $\Phi(S21)$ with respect to F, one finds that Φ_{max} occurs at

$$F = F_Z / 2$$
 (the same frequency as F_M) (34a)

$$\Phi_{\max} = 2 \cdot \arctan \frac{F_Z}{2}, \qquad (34b)$$

or

and

$$F_Z = 2 \cdot \tan \frac{\Phi_{\text{max}}}{2} \,. \tag{35}$$

As Φ_{max} is measurable, it gives another way to determine F_Z . But, it is suitable only if F_Z is quite small. Formulas (24), (30) and (35) show three ways to determine F_Z . Formula (30) is the best. The rest can be used for cross checks.

Now let's find the deviation of the Q value. From (18) and (22), one can find the 3dB bandwidth. Considering only the factors depending on frequency F in (18), the frequency at the edge of bandwidth is:

$$\frac{1 + (F + F_z)^2}{1 + F^2} = \frac{1}{2} \cdot \frac{1 + (F_{\max} + F_z)^2}{1 + F_{\max}^2}.$$
(36)

This equation can be simplified as:

$$\frac{2FF_z + F_z^2}{1 + F^2} = p, \quad \text{or} \quad pF^2 - 2FF_z - (F_z^2 - p) = 0 \quad (37a)$$

$$p = \frac{1}{2} \cdot \frac{1 + (F_{\max} + F_z)^2}{1 + F_{\max}^2} - 1.$$
 (37b)

It turns out:

$$F_{3,4} = \frac{F_z \pm \sqrt{F_z^2 + p(F_z^2 - p)}}{p}.$$
(38)

The normalized bandwidth is

$$F_3 - F_4 = \frac{2\sqrt{F_z^2 + p(F_z^2 - p)}}{p}.$$
(39)

The real Q is that in the case of no crosstalk, i.e. $F_z >>1$, $p = (F_z^2 - 1)/2$, $F_3 - F_4 = 2$. The case $F_3 - F_4 > 2$ implies that Q decreases. Therefore, the ratio of the real Q and the measured Q is:

$$G_q = \frac{\sqrt{F_z^2 + p(F_z^2 - p)}}{p}.$$
 (40)

Substituting (21),(30) and (38), after algebraic manipulation one obtains

$$G_q = \sqrt{2\left(\frac{C^2 - 1}{C^2 - 2}\right)^2 - 1}.$$
(41)

Thus the real loaded Q is

$$Q_L = G_q \ Q_{tested} \ . \tag{42}$$

RESULTS OF MEASUREMENT

We tested a superconducting cavity made of niobium, with both launcher and pickup of different sizes. Also employed was a probe C (see Fig.1), which is an axial semi-rigid cable with the tip stripped, such that it forms an antenna. The sizes of the probes are shown in Table 1. Selected measured data are in Table 2. Also shown are the deduced data calculated by means of formulas derived above. Table 3 summarizes the final coupling coefficients and the corresponding external Q. Figs.13-16 show the curves.

Probe A (Launcher)			Probe B (Pick-up)		
ID	Length (inch)	Length (mm)	ID	Length (inch)	Length (mm)
A 1	2 651	67.34	B1	2 184	55.47
Alb	2.532	64 31	B2	1 934	49.12
Alc	2.413	61.29	B3	1.684	42.77
Ald	2.248	57.10	B4	1.434	36.42
A2	2.247	57.07			
A3	2.172	55.17			

Table 1 Sizes of probes

Table 2 Selected measured data and deduced data

Port 1 probe	Ala	Ala	С	Alb	A1c	Ald	B2
Port 2_probe	C	B1	B1	C	C	C	C
f_0 (max)	1299.825	1299.868	1299.812	1299.818	1299.812	1299.812	1299.812
	272	750	716	750	500	500	500
QL	6439.8		6520.3	6432.5	6453.4	6444.8	6492.4
$f_{z}(min)$	1298.375	1299.646		1298.337	1298.131	1298.362	1298.071
	000	875		500	250	500	875
f _M	1299.100	1299.757		1299.078	1298.971	1299.087	1298.942
	640	810		125	875	500	187
$S21(f_0) dB$	-46.864	-68.449	-54.532	-49.234	-51.636	-54.954	-59.827
$S21(f_Z) dB$	-97.504	-76.65		-93.084	-98.164	-98.142	-106.02
$S21(f_M) dB$	-70.001	-73.034		-72.544	-76.28	-78.122	-84.548
φmax		51.4					
$S(f_0)$ - $S(fm)$	23.137	4.585		23.31	24.644	23.168	24.721
parameter C	14.350	1.695		14.639	17.069	14.401	17.221
Fz	14.280	1.105		14.570	17.010	14.332	17.163
f ₀ -f _Z	1.45000	0.22188		1.48125	1.68125	1.45000	1.74062

$\delta z = 2(f_0 - f_Z)/2$	2.231E-3	3.413E-4		2.279E-3	2.587E-3	2.231E-3	2.678E-3
D _{so}	4.228E-2	3.714E0	0	4.063E-2	2.986E-2	4.198E-2	2.934E-2
S21 ₀ -D _{so}	-46.906	-72.163	-54.532	-49.275	-51.666	-54.996	-59.856
$\beta_1 \beta_2$	5.097E-6	1.519E-8	8.805E-7	2.954E-6	1.704E-6	7.913E-7	2.584E-7
Gq	1.01E0	2.862E0	1E0	1.009E0	1.007E0	1.01E0	1.007E0
QL	6502.8		6521.6	6493.0	6497.9	6507.4	6536.4
Δf	6.863E-3	5.809E-2		6.728E-3	5.77E-3	6.839E-3	5.719E-3
f_0 (corrected)	1,299.818	1,299.811		1,299.812	1,299.807	1,299.806	1,299.807

Table 3. Coupling coefficients and external Q

Probe		Coup	ling Coeff	icient	External Q Qex		
ID	T.date	Jul. 16	Aug.2	Aug.29	Jul. 16	Aug.2	Aug.29
A1	β_{A1}	2.973e-4	3.089E-4	2.965e-4	2.22E7	2.01E7	2.20E7
A1b	β_{A1b}			1.718e-4			3.79E7
Alc	β_{A1c}			9.913e-5			6.57E7
Ald	β_{A1d}			4.603e-5			1.41E8
A2	β_{A2}		4.282E-5			1.45E8	
A3	β_{A3}		2.956E-5			2.10E8	
B1	β_{B1}	5.348e-5	5.144E-5	5.123e-5	1.23E8	1.21E8	1.27E8
	-		4.401E-5			1.41E8	
B2	β_{B2}	1.316e-5		1.503e-5	5.02E8		4.33E8
B3	β_{B3}	2.377e-6			2.78E9		
B4	β_{B4}	3.744e-7	2.539E-7		1.76E10	2.44E10	
C	$\beta_{\rm C}$	2.090e-2	3.038E-2	1.719e-2	3.61E5	2.00E5	3.79E5
	-		2.665E-2			2.33E7	



 $Fig. 13. \ Dependence \ of \ coupling \ coefficient \ on \ launcher \ length$



Fig.14. Dependence of coupling coefficient on pick-up length

Fig.15. Dependence of external Q on launcher length



Fig.16. Dependence of external Q on pick-up length

The coupling coefficient and external Q are the most important parameters that need to be determined. From the description above, one can correct the resonant frequency from (23) and the Q value from (41) and (42).

The curves show good results within a reasonable error and thus justify the method in this note.

ACKNOWLEDGEMENT

This work is supported by the project of "Electron Cooling" headed by Ilan Ben-Zvi, in cooperation with AES. The cavity experiment is managed by Triveni Srinivasan-Rao and Mike Cole of BNL and AES, respectively. Mike Cole participated the cavity measurements in all aspect.

References

[1] Mike Cole, private communication

- [2] Yongxiang Zhao, "SC cavity cross-talk measurement and analysis", presented in Superconducting Cavity Gun Meeting, Advanced Energy Systems, NY, July 22, 2002.
- [3] Yongxiang Zhao, "Test report for SC cavity", Aug. 9, 2002.
- [4] Yongxiang Zhao, "The report for the SC cavity measurement", Sept.5, 2002.