

Quad Trim Winding Interconnection on 21Q40 Half Cell Wiring Diagrams

R. F. Lambiase

June 2004

Collider Accelerator Department
Brookhaven National Laboratory

U.S. Department of Energy

USDOE Office of Science (SC)

Notice: This technical note has been authored by employees of Brookhaven Science Associates, LLC under Contract No.DE-AC02-98CH10886 with the U.S. Department of Energy. The publisher by accepting the technical note for publication acknowledges that the United States Government retains a non-exclusive, paid-up, irrevocable, world-wide license to publish or reproduce the published form of this technical note, or allow others to do so, for United States Government purposes.

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or any third party's use or the results of such use of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof or its contractors or subcontractors. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

Quad Trim Winding Interconnection on 21Q40 Half Cell Wiring Diagrams

BNL/SNS Technical Note
No. 136
R. F. Lambiase
June 2, 2004

Collider-Accelerator Department
Brookhaven National Laboratory
Upton, New York 11973

Overview

I have reviewed seven wiring diagrams, all relating to 21Q40 installations. The drawings are 5300 074, 081, 082, 086, 099, 100, and 129. The trim wiring on all these drawings is correct for the depiction of the winding direction shown on the drawings.

Methods

SNS Tech Note 119 gives us the pole definition for “A” polarity for a normal quad when looking at the lead end of the magnet. This is shown in Figure 1. The lead end is defined by where the main winding leads emerge. The terminal blocks are on the opposite end, so when viewing the terminals blocks for the trim windings, the pole definitions are mirror images. This is shown in Figure 2.




Figure 1 - Lead End, “A” Polarity





Figure 2 - Non-lead End, “A” Polarity

With a correct interconnection, a positive current entering terminal 1 will produce an “A” Polarity. I checked the drawings by marking up the relevant portion of each. The results are attached as a reference.


Limitations

For this drawing update, no coil drawings were changed. It is assumed by the drawing set that the person attaching the trim leads to the terminals will know how to do so without identification of the leads on the coil winding diagram.




Dwg 5300074

TRIM COILS WIRING DIAGRAM OF QUAD 21Q40
VIEW TOWARD DIPOLE




Dwg 5300081

TRIM COILS WIRING DIAGRAM OF QUAD 21Q40
VIEW TOWARD DIPOLE




TRIM COILS WIRING DIAGRAM OF QUAD 21Q40
VIEW TOWARD DIPOLE




Dwg 5300086


TRIM COILS WIRING DIAGRAM OF QUAD 21Q40
VIEW TOWARD DIPOLE



TRIM COILS WIRING DIAGRAM OF QUAD 21Q40
VIEW TOWARD DIPOLE



TRIM COILS WIRING DIAGRAM OF QUAD 21Q40
VIEW TOWARD DIPOLE



Dwg 5300129

TRIM COILS WIRING DIAGRAM OF QUAD 21Q40
VIEW LOOKING AT NON-LEAD END