

BNL-105628-2014-TECH

BNL/SNS Technical Note No. 059;BNL-105628-2014-IR

Ionization Profile Monitor for SNS

R. Connolly

March 1999

Collider Accelerator Department Brookhaven National Laboratory

U.S. Department of Energy

USDOE Office of Science (SC)

Notice: This technical note has been authored by employees of Brookhaven Science Associates, LLC under Contract No.DE-AC02-98CH10886 with the U.S. Department of Energy. The publisher by accepting the technical note for publication acknowledges that the United States Government retains a non-exclusive, paid-up, irrevocable, world-wide license to publish or reproduce the published form of this technical note, or allow others to do so, for United States Government purposes.

DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or any third party's use or the results of such use of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof or its contractors or subcontractors. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.



Ionization Profile Monitor for SNS

BNL/SNS TECHNICAL NOTE

NO. 059

R. Connolly, P. Cameron, T. J. Shea and R. Witkover

March 1, 1999

ALTERNATING GRADIENT SYNCHROTRON DEPARTMENT BROOKHAVEN NATIONAL LABORATORY UPTON, NEW YORK 11973

IONIZATION PROFILE MONITOR FOR SNS

R. Connolly, P. Cameron, T.J. Shea, and R. Witkover Brookhaven National Lab, Upton, New York 11973

INTRODUCTION

The Spallation Neutron Source (SNS) consists of a 1 GeV linear accelerator (linac), a storage ring, a target, and connecting beam-transport lines [1]. The linac delivers a 1.04-ms-long macropulse which is chopped into about 10^3 pulses of 10^{11} protons each. As this pulse enters the storage ring from the linac each chopped pulse lines up longitudinally with all proceeding pulses forming a single bunch about 550ns long and a 250-ns gap. During the accumulation process the transverse profiles of the beam will be measured with ionization profile monitors (IPM's).

An IPM measures the distribution of electrons freed by ionizing collisions of the beam with residual gas in the beamline [2,3]. The electrons are swept from the beamline by a transverse electric field, amplified by a microchannel plate (MCP), and collected by a circuit board with strip anodes oriented parallel to the beam axis. A uniform magnetic field, parallel to the sweep electric field, counters the defocusing effects of space charge and recoil momentum [4]. During each bunch passage the charge pulses are integrated, amplified, and digitized for display as a profile histogram.

The purpose of this paper is to evaluate the performance requirements the SNS will demand of an IPM and to propose a development path to meet these requirements. In the following sections we discuss electron focusing, proposed detector design, anticipated signal sizes, and requirements for the electronics. It is logical at this time to propose that the SNS IPM will be similar to those used in RHIC. However only a prototype has been tested for RHIC. Beam tests of the RHIC IPM will begin in the summer of 1999 and these tests will provide valuable information for the design of IPM's for SNS.

All model calculations assume a full beam emittance of 120π mm*mrad and β =10m giving a beam 5 σ radius of 35mm. Since the chopped linac pulses are injected around the circumference of a ring in transverse phase space, the transverse profiles of the accumulated bunch in the ring are approximately flat with peaks at both ends [5]. The calculations model the incoming chopped pulses with Gaussian transverse profiles and the full beam with a rectangular transverse profile.

ELECTRON KINEMATICS

The density of ionization events in the residual gas is a map of the beam distribution. An IPM sweeps the electrons formed in these ionizing collisions from the beamline with a transverse electric field and measures their distribution. While the electrons are being swept from the beampipe they can move parallel, but not perpendicular, to the collectors anodes. However, the ionizing event imparts momentum to each electron, and the electron is released in the space-charge electric field of the beam. The impulse from the collision and the transverse component of the space-charge electric field can move the electron perpendicular to the collector channels resulting in a measured profile which is different than the beam's. To counter these effects the detector is placed in a magnetic field oriented parallel to the applied electric field.

The space-charge electric field of the beam bunch is a radial field whose strength increases with radius to a maximum near the beam's edge. Each electron is released in an electric field with a component parallel to the applied sweep field (which causes no error) and a component perpendicular to it. Without magnetic confinement the perpendicular component, E_{\perp} , accelerates the electron across collector channels. With the magnetic field the electron is in perpendicular electric and magnetic fields, and it will drift with a cycloidal motion in the direction defined by the vector **E** × **B** (parallel to the collector anodes) with a velocity of,

$$\mathbf{v} = (\mathbf{E} \times \mathbf{B})/\mathbf{B}^2 \tag{1}$$

and a gyration radius of,

$$\mathbf{R} = \mathbf{Em}/\mathbf{qB}^2.$$
 (2)

A bunch containing $2x10^{14}$ protons within a radius of 35mm (β =10m) will have a maximum space-charge field of about 10^5 V/m. In a magnetic field of 0.1 T, an electron in the maximum transverse field will travel parallel to the collector channels at 10^6 m/s with a gyration radius of <70 µm. Profile broadening by the space-charge field is insignificant. By the same mechanism transverse components in the sweep electric field will not cause measurement errors. A sweep field of 150kV/m will remove electrons is under 10ns so the maximum longitudinal drift is less than 1.0cm.

The second defocusing effect is the momentum impulse from the ionizing collision. The energy spectrum of recoil electrons extends to 3.0 MeV in SNS but over 95% of them have energies <500 eV [6]. In a magnetic field the electron trajectory is a helix whose axis is parallel to the magnetic field and whose radius varies linearly with the transverse component of momentum. A 0.1T field confines a 500 eV electron to a Larmor radius of <0.8 mm which is about 2% of the beam radius. By placing the detector in a field of 0.1T most of the electrons are collected on the anodes over which they are formed.

DETECTOR DESIGN

Figure 1 shows the RHIC IPM. The detector, three electrical feedthroughs, and a quartz window (not shown) are mounted on a single 10" conflat flange. This flange is mounted on a side port of the detector vacuum chamber made from 100x150mm stainless steel tubing. The vacuum chamber is mounted in the gap of a 'C' dipole permanent magnet.



Figure 1. Cross section of RHIC IPM. View is looking down the beamline.

Two rectangular brackets hold the collector board and a double-plate MCP amplifier on one side of the beam, and a sweep-field electrode and secondary electron suppression grid on the other side. The collector board is 0.625 mm-thick alumina, plated with gold, and etched with 64 collection anodes. At one end of each anode a plated through hole conducts the charge to a trace on the back which takes the signal to the edge of the board. A wire connects each trace to a pin on a vacuum feedthrough.

For RHIC we use permanent-magnet dipoles producing fields of 0.19T. Data taken with the RHIC prototype IPM indicate that 0.1T is adequate for accurate imaging. However the magnets were built with strontium-ferrite bricks (1"x4"x6") and integral multiples of 14 were required for field uniformity. A 14-brick magnet was too weak so we used 28 bricks. In addition to the detector magnet, a second magnet is mounted 0.5 m from the first with its field oriented opposite to the first. The beam experiences a slight transverse offset passing through the chamber but emerges with the direction of its trajectory unchanged. For SNS we will use three magnets for each IPM with the two end magnets causing half of the beam deflection as the center IPM magnet.

The SNS beam has a full-beam (5σ) emittance of 120π mm*mrad so the design beam diameter at β =10m is 35mm. Collimators in the beamline restrict the full-beam emittance to 180π mm*mrad so the maximum beam size where β =10m is 84mm diameter. Using these beam sizes a detector with 64 collector channels spaced 1.4 mm apart covering a diameter of 90mm would be appropriate. The full beam would produce signals in 50 channels and all beam, including halo, would cover 60 channels. A clear aperture of twice the maximum beam diameter will be used to minimize background from scattered beam.

SIGNAL SIZE

It is expected that the residual gas in SNS will be about two thirds H_2 and one third CO [7]. The ionization rate produced by a proton traveling through a gas decreases with increasing energy to a minimum at about 1 GeV [8]. The ionization rate consists of primary ionization rates and secondary events caused by the energetic recoil electrons. Figure 2, copied from ref. 8, shows the approximate relationship between primary ionization rates and atomic number for gases at atmospheric pressure for a minimum ionizing particle.



Figure 2. The number of primary ionizing events produced by 1 GeV protons per cm at one atmosphere pressure plotted vs. the atomic number of the gas. This plot is copied from ref. 9.

The measured primary ionization rate in H_2 is 5.2 ion pairs/cm and, from this graph, we estimate the ionization rate in CO to be 22 i.p./cm. Therefore a 1 GeV proton in this mixture of gases at one atmosphere pressure produces 10.7 i.p./cm. Each bunch passage past the detector will create N detected electrons where,

N=[1.4x10-2][Pressure(torr)][Number of protons][Detector length(cm)][Detection efficiency]. (3)

An MCP detection efficiency is about 36% for electrons with energies of a few keV. Galileo produces rectangular MCP with active areas of 74x93mm [9]. A simple detector configuration would be to cover the 900mm-wide collector with a single MCP giving a detector length of 74mm. The number of detected electrons then is,

 $N=[3.7x10^{-2}][Pressure(torr)][Number of protons]$ (4)

With a beamline pressure of $5x10^{-9}$ torr an injected bunch at half the design intensity would produce 10 electrons and the full beam intensity of $2x10^{14}$ protons would produce $3.7x10^4$.

It is not feasible to build a system with this large of dynamic range. Primarily the IPM will be used to measure the profile toward the end of the machine cycle. The MCP will remain biased and the sweep field will be switched on and off to record the interesting portion of the machine cycle. For machine studies of the injection process the chamber pressure will be raised by a controlled leak with feedback. The design injected beam bunch is 10^{11} protons with an emittance of 0.14π mm*mrad. If the IPM is located at β =10m the rms width of the beam is 1.2mm. This will deliver signal to only 3-4 IPM channels. This will not provide an emittance measurement but will allow monitoring of phase-space filling.

A smoke ring distribution in phase space results in a profile which is approximately rectangular. With 3.7×10^4 electrons covering 50 channels each channel receives about 740. If a single MCP is operated with a gain of 10^4 , then each channel would deliver a current of about 2 μ A during the beam pulse. These currents will be brought out of the tunnel on coaxial cable and amplified and digitized.

ELECTRONICS

Microchannel plate

If a MCP is used as the primary amplifier and the profile of every bunch is to measured, the maximum detector output charge per beam bunch is limited by the MCP strip current. When a MCP is used as a dc current amplifier, the gain is constant until the output current exceeds approximately 10% of the strip current [10]. A 74x93 mm extended dynamic range MCP has a strip current of 325-1500 μ A when biased at 1 kV. If we assume a current of 1.0mA per plate this gives an upper limit on the signal current density of 0.014 μ A/mm². Each channel is 104 mm² so the output current per channel should kept below 1.5 μ A. With a bunch frequency of 1.19MHz this gives a maximum charge per pulse per channel of 1.3 pC. The example worked out in the previous section gave 1.2pC/bunch.

A MCP has a finite lifetime caused by electron scrubbing of the channel walls near the plate output. The gain gradually drops, falling to about 90% of new after the plate has delivered about 0.1C/cm². Each collector channel has an area of about 1 cm² so the lifetime of the plate will be approximately 8×10^{10} measured bunches. If 100 turns are measured each machine cycle the plate change interval will be 8×10^{8} cycles. Also the finite lifetime requires the plate bias voltage be interlocked to the measurement system so it is only on during measurements.

Charge integration

To keep the MCP from saturating during the machine cycle the charge per channel for each beam pulse has to be limited to about 1.0 pC. For a reasonable dynamic range across the profile the electronics have to be sensitive to integrated charges of 10 fC. Such low quantities of charge are difficult to measure with gated integrators. For instance the Burr-Brown ACF2101 gated integrator has an integration capacitor of 100pF (10mV/pC) and typical charge injection at reset and read of 100-200 fC [11]. There are ways to compensate for charge injection or to subtract it from the signal, but we do not know if it has been done well enough to achieve the required charge sensitivity. If single-turn profiles are not needed, a profile from many turns can be gathered in a single integration cycle thus reducing the charge-injection problem.

During the machine cycle the space charge electric field of the beam goes from zero to $>10^{5}$ V/m. The field from the proton beam will keep electrons in the beam pipe during bunch passage. If electrons are to be extracted during the bunch passage the sweep field has to be larger than the maximum space-charge field. Since the maximum space charge field is not an exactly predictable quantity some overhead will have to be allowed. Probably an applied field of 150kV/m would be adequate for all situations. This is achievable with 15kV applied over a 10cm gap.

Another possible data-collection method is to amplify the signals and digitize them many times during the bunch passage. The VME digitizers developed for RHIC could digitize the profiles five times during one bunch passage. Integration over the beam pulse would be done by adding up the several profiles collected during one bunch passage. This is the proposal for the beam position monitor system [12].

DISCUSSION

The profile of the SNS beam can be measured with a detector similar to that used for RHIC. Two challenges are to integrate the very small charge signals over the 550ns bunch and to follow three decades of beam current. The charge integration might be done with either low-noise gated integrators or with fast digitizers. The large beam dynamic range will be handled by designing a system to measure beam profiles of the full beam and creating a local pressure bump with a small, controlled leak of an easily pumped heavy gas such as CO for machine studies earlier in the cycle.

Some care will be required when locating the detector. MCP's are sensitive to any radiation which produces secondary electrons. The detector cannot be near any beamline location where beam loss is likely such as scrapers or collimators. All of the beamline components of the detector have to lie outside of the maximum beam envelope defined by the collimator. With the intense beam of SNS it would be a good idea to place an annular shield in the beamline upstream from the IPM to keep beam-generated x-rays from causing spurious signals.

One final concern is the operation of MCP's in a high radiation environment. Any radiation which produces secondary electrons can excite a channel. An MCP is sensitive to x-rays with wavelengths as short as 0.12Å. It is possible that radiation will cause a large background with the channel plate biased. This has to be investigated.

REFERENCES

- 1. SNS Design Manual
- W.H. DeLuca, "Beam Detection Using Residual Gas Ionization", IEEE Trans. Nucl. Sci., <u>NS-16</u>, 813 (1969).
- A.N. Stillman, et. al., "An Ultrahigh Vacuum Beam Profile Monitor", Rev. Sci. Instrum., <u>63</u> (6) (1992)
- 4. R. Connolly, P. Cameron, W. Ryan, T.J. Shea, R. Sikora, and N. Tsoupas, "A Prototype

Ionization Profile Monitor for RHIC", Proceedings of the 1997 Particle Accelerator Conference.

- 5. A.U. Luccio, D. Maletic and F.W. Jones, "Injection and RF Capture in the NSNS; A Numerical Simulation in 6 Dimensions," BNL/NSNS Technical Note No. 004, Brookhaven National Laboratory, Upton, NY.
- 6. J. D. Jackson, "Classical Electrodynamics," p. 430, John Wiley and Sons, New York, 1962.
- 7. Dick Hseuh, Private communication.
- 8. F. Sauli, "Principles of Operation of Multiwire Proportional and Drift Chambers," CERN 77-09 (1977).
- 9. Galileo Electro-Optics Corp., Sturbridge, MA 01566.
- 10. J. Wiza, "Microchannel Plate Detectors", Nucl. Instr. and Meth. <u>162</u> (1979) 587.
- 11. Burr-Brown Corp., Tucson, AZ 85734.
- 12. T.J. Shea, P.R. Cameron, and W.A. Ryan, "Design Study of Position Monitoring in the SNS Ring", Brookhaven National Laboratory, Upton, NY 11973.