

Requirements for CEC POP machine protection system

I. Pinayev

February 2015

Collider Accelerator Department
Brookhaven National Laboratory

U.S. Department of Energy

USDOE Office of Science (SC), Nuclear Physics (NP) (SC-26)

Notice: This technical note has been authored by employees of Brookhaven Science Associates, LLC under Contract No. DE-SC0012704 with the U.S. Department of Energy. The publisher by accepting the technical note for publication acknowledges that the United States Government retains a non-exclusive, paid-up, irrevocable, world-wide license to publish or reproduce the published form of this technical note, or allow others to do so, for United States Government purposes.


DISCLAIMER

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, nor any of their contractors, subcontractors, or their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or any third party's use or the results of such use of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or any agency thereof or its contractors or subcontractors. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.

C-A/AP/537
February 2015

Requirements for CEC POP Machine Protection System

I. Pinayev



**Collider-Accelerator Department
Brookhaven National Laboratory
Upton, NY 11973**

**U.S. Department of Energy
Office of Science, Office of Nuclear Physics**

Notice: This document has been authorized by employees of Brookhaven Science Associates, LLC under Contract No. DE-SC0012704 with the U.S. Department of Energy. The United States Government retains a non-exclusive, paid-up, irrevocable, world-wide license to publish or reproduce the published form of this document, or allow others to do so, for United States Government purposes.

REQUIREMENTS FOR CEC POP MACHINE PROTECTION SYSTEM

The machine protection system should prevent damage of a vacuum chamber with missteered electron beam. For the evaluation purposes we will utilise the following beam electron parameters:

Energy	– 22 MeV
Maximal bunch charge	– 5 nC
Maximal repetition rate	– 78 kHz
Normalized emittance	– 5 mm mrad
Minimal β -function	– 1 m

Vacuum system is build of steel, for which total stopping power of the 22 MeV electron beam is about 2.8 MeV g/cm² [1]. For the steel density of 7.8 g/cm³ specific losses (per unit of length) will be 21.8 MeV/cm. For 2 MeV electron beam stopping power is 1.4 MeV g/cm² [1] and losses will be 10.9 MeV/cm.

Maximal density of the electron current is defined by maximal current of $I_{max}=390 \mu\text{A}$ and minimal beam size, which is assumed equal in both planes. At $\gamma=43$ the r.m.s. emittance will be 120 nm and r.m.s. beam size in each plane will be $\sigma=340 \mu$. The peak current density for Gaussian beam is equal to $I_{max}/2\pi\sigma^2 = 0.054 \text{ A/cm}^2$, and volumetric power dissipation will be 1.17 MW/cm³. The volumetric heat capacity of iron is 3.5 J/cm³/K. Allowing temperature excursions up to $\Delta T=200^\circ\text{K}$ we can find the maximal reaction time of the machine protection system

$$t_{max}=3.5 \times 200 / 1.17 \times 10^6 = 600 \mu\text{s}$$

It should be noted that the obtained result does not depend either beam has normal or shallow incidence on the surface.

Now we will estimate how much current loss we can withstand in long run. Essentially is the same limit below which we can perform any beam manipulations without destroying the vacuum chamber or diagnostics equipment. Because the transverse size of the electron beam is much smaller than the loss length (1 cm for 22 MeV beam, and 2 mm for 2 MeV beam) we can consider the cylindrical geometry. Let consider the electron beam with radius r , and uniform current density $j=I/\pi r^2$. The volumetric heat dissipation will be $H=jL$, where L is specific losses in the wall. Inside the cylinder of radius $x<r$ and length l total generated heat will be $P=jL\pi x^2 l$, and these power will be dissipated through conductance $P=-2\pi x l k dT/dx$, where k is thermal conductivity (80 W/K/m for iron). Or $dT/dx=-jLx/2k$ and

$$T(x)=T_{max}-jLr^2/4k(x/r)^2 = T_{max}-IL/4\pi k(x/r)^2$$

for $x \leq r$. For $x > r$ the amount of heat generated is fixed $P=ILl$ and temperature gradient is $dT/dx=-IL/2\pi kx$. Temperature dependence on distance is logarithmic, therefore we will assume that at some distance R temperature is fixed $T=0$. Then dependence will be $T(x)=IL\ln(R/x)/2\pi k$, and temperature at beam boundary will be $T(r)=IL\ln(R/r)/2\pi k$. Now we can find T_{max} :

$$T_{max}=IL/4\pi k+IL\ln(R/r)/2\pi k= IL/4\pi k(1+1/2 \ln(R/r))$$

For R we can choose triple loss length and for r the smallest beam size. For this case

$$T_{max} \approx IL/4k$$

and assuming maximal temperature rise of 35°C the maximal allowable loss current is

$$I_{loss\ max} = 4kT_{max}/L = 4 \times 0.8 \times 35 / 21.8 \times 10^6 = 5.1 \mu A$$

REFERENCE

1. <http://physics.nist.gov/PhysRefData/Star/Text/ESTAR.html>